

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076470.D
 Acq On : 30 Jan 2023 20:30
 Operator : JC\MD
 Sample : 01333-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-11

Quant Time: Jan 31 00:46:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	386389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	665246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	595017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	229881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	239639	55.391	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.780%		
35) Dibromofluoromethane	8.177	113	201896	48.516	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.040%		
50) Toluene-d8	10.571	98	788418	51.301	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.600%		
62) 4-Bromofluorobenzene	12.853	95	312764	62.340	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 124.680%#		
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	34277	24.649	ug/l	93
20) Methylene Chloride	5.289	84	29796	6.739	ug/l	90
37) Ethyl Acetate	7.577	43	41426	8.764	ug/l	98
43) Isopropyl Acetate	8.694	43	478048	61.957	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076470.D
Acq On : 30 Jan 2023 20:30
Operator : JC\MD
Sample : 01333-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-11

Quant Time: Jan 31 00:46:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

