

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076471.D
 Acq On : 30 Jan 2023 20:54
 Operator : JC\MD
 Sample : 01352-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2-WC

Quant Time: Jan 31 00:47:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	376574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	676103	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	697047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	221333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	233069	55.276	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.560%
35) Dibromofluoromethane	8.177	113	190433	45.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.060%
50) Toluene-d8	10.571	98	752826	48.198	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.400%
62) 4-Bromofluorobenzene	12.853	95	240283	47.124	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.240%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	57851	42.739	ug/l	100
20) Methylene Chloride	5.289	84	28507	6.605	ug/l	97
37) Ethyl Acetate	7.571	43	39801	8.287	ug/l	99
43) Isopropyl Acetate	8.694	43	483254	61.626	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076471.D
Acq On : 30 Jan 2023 20:54
Operator : JC\MD
Sample : 01352-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2-WC

Quant Time: Jan 31 00:47:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

