

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
 Data File : VN080810.D
 Acq On : 30 Jan 2024 13:13
 Operator : JC\MD
 Sample : P1290-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-195-GW-540-542

Quant Time: Jan 31 02:09:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	90298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	223543	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	223960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	72632	46.495	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.000%
35) Dibromofluoromethane	8.171	113	58221	42.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.140%
50) Toluene-d8	10.571	98	238151	44.490	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.980%
62) 4-Bromofluorobenzene	12.853	95	89627	48.049	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.100%
Target Compounds						
16) Acetone	4.436	43	2888	5.098	ug/l	Qvalue # 62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
Data File : VN080810.D
Acq On : 30 Jan 2024 13:13
Operator : JC\MD
Sample : P1290-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-195-GW-540-542

Quant Time: Jan 31 02:09:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

