

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
 Data File : VN080811.D
 Acq On : 30 Jan 2024 13:37
 Operator : JC\MD
 Sample : P1288-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-191-TB-20240126

Quant Time: Jan 31 02:09:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	89501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	214824	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	227767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88622	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	81047	52.344	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.680%
35) Dibromofluoromethane	8.165	113	62255	47.367	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.740%
50) Toluene-d8	10.565	98	262861	51.100	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.200%
62) 4-Bromofluorobenzene	12.853	95	99071	55.267	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
Data File : VN080811.D
Acq On : 30 Jan 2024 13:37
Operator : JC\MD
Sample : P1288-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-191-TB-20240126

Quant Time: Jan 31 02:09:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

