

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
 Data File : VN080815.D
 Acq On : 30 Jan 2024 15:13
 Operator : JC\MD
 Sample : P1288-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-191-GW-583-585

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/31/2024
 Supervised By : Semsettin Yesilyurt 01/31/2024

Quant Time: Jan 31 02:11:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	85107	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	205908	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	220196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84588	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	76453	51.927	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.860%			
35) Dibromofluoromethane	8.165	113	57908	45.968	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.940%			
50) Toluene-d8	10.571	98	241677	49.016	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.040%			
62) 4-Bromofluorobenzene	12.847	95	93636	54.497	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 109.000%			
Target Compounds					Qvalue	
16) Acetone	4.418	43	3338m	6.251	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
Data File : VN080815.D
Acq On : 30 Jan 2024 15:13
Operator : JC\MD
Sample : P1288-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-191-GW-583-585

Quant Time: Jan 31 02:11:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/31/2024
Supervised By :Semsettin Yesilyurt 01/31/2024

