

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
 Data File : VN080816.D
 Acq On : 30 Jan 2024 15:36
 Operator : JC\MD
 Sample : P1288-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-191-GW-598-600

Manual Integrations
 APPROVED

Reviewed By :John Carlone 01/31/2024
 Supervised By :Semsettin Yesilyurt 01/31/2024

Quant Time: Jan 31 02:12:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	95631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	230655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	234483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	78292	47.324	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.640%		
35) Dibromofluoromethane	8.165	113	58221	41.258	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	82.520%		
50) Toluene-d8	10.565	98	248931	45.070	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.140%		
62) 4-Bromofluorobenzene	12.847	95	95260	49.494	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.980%		
Target Compounds					Qvalue	
16) Acetone	4.424	43	3272m	5.453	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
Data File : VN080816.D
Acq On : 30 Jan 2024 15:36
Operator : JC\MD
Sample : P1288-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-191-GW-598-600

Quant Time: Jan 31 02:12:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/31/2024
Supervised By :Semsettin Yesilyurt 01/31/2024

