

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
 Data File : VN080818.D
 Acq On : 30 Jan 2024 16:48
 Operator : JC\MD
 Sample : P1265-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 31 02:13:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	94711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	224771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	229346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88972	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	74967	45.754	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.500%
35) Dibromofluoromethane	8.165	113	59560	43.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.620%
50) Toluene-d8	10.565	98	245237	45.564	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.120%
62) 4-Bromofluorobenzene	12.853	95	93572	49.890	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
Data File : VN080818.D
Acq On : 30 Jan 2024 16:48
Operator : JC\MD
Sample : P1265-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 31 02:13:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

