

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
 Data File : VN080826.D
 Acq On : 30 Jan 2024 19:58
 Operator : JC\MD
 Sample : P1265-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 31 02:14:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	85069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	205227	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	213208	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78613	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	71741	48.748	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.500%
35) Dibromofluoromethane	8.165	113	55867	44.495	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.980%
50) Toluene-d8	10.571	98	227507	46.295	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.600%
62) 4-Bromofluorobenzene	12.847	95	85331	49.828	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
Data File : VN080826.D
Acq On : 30 Jan 2024 19:58
Operator : JC\MD
Sample : P1265-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 31 02:14:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

