

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
 Data File : VN080827.D
 Acq On : 30 Jan 2024 20:22
 Operator : JC\MD
 Sample : P1270-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-LIQUID-20240126

Quant Time: Jan 31 02:14:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	84449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	205573	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	206987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	73443	50.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.540%
35) Dibromofluoromethane	8.165	113	58859	46.799	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.600%
50) Toluene-d8	10.565	98	230267	46.778	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.560%
62) 4-Bromofluorobenzene	12.853	95	86129	50.210	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.420%
Target Compounds						
16) Acetone	4.424	43	12578	23.739	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013024\
Data File : VN080827.D
Acq On : 30 Jan 2024 20:22
Operator : JC\MD
Sample : P1270-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-LIQUID-20240126

Quant Time: Jan 31 02:14:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

