

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN013025\  
 Data File : VN085578.D  
 Acq On : 30 Jan 2025 14:32  
 Operator : JC\MD  
 Sample : VN0130WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0130WBL01

Quant Time: Jan 31 02:34:41 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011425W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jan 15 02:16:08 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon           | Response | Conc       | Units | Dev(Min) |
|-----------------------------|--------|----------------|----------|------------|-------|----------|
| -----                       |        |                |          |            |       |          |
| Internal Standards          |        |                |          |            |       |          |
| 1) Pentafluorobenzene       | 8.224  | 168            | 169481   | 50.000     | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100  | 114            | 315219   | 50.000     | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.865 | 117            | 275647   | 50.000     | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.788 | 152            | 110956   | 50.000     | ug/l  | 0.00     |
| System Monitoring Compounds |        |                |          |            |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.577  | 65             | 154457   | 56.458     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 74 - 125 | Recovery | = 112.920% |       |          |
| 35) Dibromofluoromethane    | 8.165  | 113            | 117387   | 53.679     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 75 - 124 | Recovery | = 107.360% |       |          |
| 50) Toluene-d8              | 10.565 | 98             | 389115   | 50.080     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 86 - 113 | Recovery | = 100.160% |       |          |
| 62) 4-Bromofluorobenzene    | 12.847 | 95             | 120418   | 45.306     | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range 77 - 121 | Recovery | = 90.620%  |       |          |

Target Compounds Qvalue

-----  
 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN013025\  
Data File : VN085578.D  
Acq On : 30 Jan 2025 14:32  
Operator : JC\MD  
Sample : VN0130WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0130WBL01

Quant Time: Jan 31 02:34:41 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011425W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jan 15 02:16:08 2025  
Response via : Initial Calibration

