

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013025\
 Data File : VN085581.D
 Acq On : 30 Jan 2025 15:54
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 31 02:36:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	312142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150035	54.983	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.960%		
35) Dibromofluoromethane	8.165	113	112634	52.013	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.020%		
50) Toluene-d8	10.565	98	382724	49.743	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.480%		
62) 4-Bromofluorobenzene	12.847	95	123710	47.004	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 94.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013025\
Data File : VN085581.D
Acq On : 30 Jan 2025 15:54
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 31 02:36:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

