

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013120\
 Data File : VN059945.D
 Acq On : 31 Jan 2020 14:11
 Operator : JC/MD
 Sample : L1325-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleID :
 200029

Quant Time: Feb 03 03:08:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	302999	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	281032	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	125195	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	109264	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
35) Dibromofluoromethane	7.57	113	91197	50.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.94%	
50) Toluene-d8	10.08	98	365172	52.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.32%	
62) 4-Bromofluorobenzene	12.40	95	129335	51.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.50%	
Target Compounds						
43) Isopropyl Acetate	8.14	43	49961	11.737	ug/l #	91
51) 4-Methyl-2-Pentanone	9.97	43	3539	1.438	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013120\
 Data File : VN059945.D
 Acq On : 31 Jan 2020 14:11
 Operator : JC/MD
 Sample : L1325-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 200029

Quant Time: Feb 03 03:08:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

