

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013120\
 Data File : VN059947.D
 Acq On : 31 Jan 2020 14:56
 Operator : JC/MD
 Sample : L1325-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11976

Quant Time: Feb 03 03:16:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	196302	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	302189	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	280035	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	130544	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	111000	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
35) Dibromofluoromethane	7.57	113	91101	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
50) Toluene-d8	10.08	98	363376	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.08%	
62) 4-Bromofluorobenzene	12.40	95	131224	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
Target Compounds						
43) Isopropyl Acetate	8.14	43	48426	11.407	ug/l #	92
51) 4-Methyl-2-Pentanone	9.97	43	3175	1.294	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013120\
Data File : VN059947.D
Acq On : 31 Jan 2020 14:56
Operator : JC/MD
Sample : L1325-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11976

Quant Time: Feb 03 03:16:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

