

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN013120\
 Data File : VN059938.D
 Acq On : 31 Jan 2020 10:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 01:57:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.534	0.519	2.8	84	0.00
3 P	Chloromethane	0.572	0.495	13.5	83	0.00
4 C	Vinyl Chloride	0.655	0.647	1.2#	89	0.00
5 T	Bromomethane	0.400	0.388	3.0	92	0.00
6 T	Chloroethane	0.310	0.305	1.6	95	0.00
7 T	Trichlorofluoromethane	0.787	0.775	1.5	93	0.00
8 T	Diethyl Ether	0.276	0.283	-2.5	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.438	0.461	-5.3	99	0.00
10 T	Methyl Iodide	0.708	0.617	12.9	83	0.00
11 T	Tert butyl alcohol	0.093	0.075	19.4	76	-0.01
12 CM	1,1-Dichloroethene	0.457	0.434	5.0#	92	0.00
13 T	Acrolein	0.066	0.041	37.9#	62	0.00
14 T	Allyl chloride	0.617	0.609	1.3	94	0.00
15 T	Acrylonitrile	0.213	0.205	3.8	89	0.00
16 T	Acetone	0.224	0.252	-12.5	124	0.00
17 T	Carbon Disulfide	1.303	1.076	17.4	79	0.00
18 T	Methyl Acetate	0.649	0.634	2.3	94	0.00
19 T	Methyl tert-butyl Ether	1.491	1.516	-1.7	98	0.00
20 T	Methylene Chloride	0.557	0.497	10.8	96	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.464	4.5	92	0.00
22 T	Diisopropyl ether	1.397	1.431	-2.4	98	0.00
23 T	Vinyl Acetate	1.082	1.081	0.1	101	0.00
24 P	1,1-Dichloroethane	0.854	0.861	-0.8	98	0.00
25 T	2-Butanone	0.314	0.312	0.6	98	0.00
26 T	2,2-Dichloropropane	0.821	0.833	-1.5	100	0.00
27 T	cis-1,2-Dichloroethene	0.559	0.565	-1.1	97	0.00
28 T	Bromochloromethane	0.341	0.349	-2.3	100	0.00
29 T	Tetrahydrofuran	0.195	0.172	11.8	85	0.00
30 C	Chloroform	0.885	0.925	-4.5#	100	0.00
31 T	Cyclohexane	0.822	0.716	12.9	87	0.00
32 T	1,1,1-Trichloroethane	0.818	0.830	-1.5	96	0.00
33 S	1,2-Dichloroethane-d4	0.555	0.544	2.0	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.298	0.301	-1.0	93	0.00
36 T	1,1-Dichloropropene	0.435	0.436	-0.2	94	0.00
37 T	Ethyl Acetate	0.408	0.375	8.1	90	0.00
38 T	Carbon Tetrachloride	0.474	0.486	-2.5	98	0.00
39 T	Methylcyclohexane	0.537	0.511	4.8	90	0.00
40 TM	Benzene	1.299	1.292	0.5	96	0.00
41 T	Methacrylonitrile	0.201	0.175	12.9	84	-0.01
42 TM	1,2-Dichloroethane	0.440	0.471	-7.0	102	0.00
43 T	Isopropyl Acetate	0.702	0.663	5.6	90	0.00
44 TM	Trichloroethene	0.354	0.378	-6.8	98	0.00
45 C	1,2-Dichloropropane	0.321	0.337	-5.0#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN013120\
 Data File : VN059938.D
 Acq On : 31 Jan 2020 10:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 03 01:57:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.227	0.234	-3.1	97	0.00
47 T	Bromodichloromethane	0.457	0.493	-7.9	102	0.00
48 T	Methyl methacrylate	0.314	0.293	6.7	88	0.00
49 T	1,4-Dioxane	0.006	0.005	16.7	79	0.00
50 S	Toluene-d8	1.155	1.145	0.9	92	0.00
51 T	4-Methyl-2-Pentanone	0.406	0.384	5.4	89	0.00
52 CM	Toluene	0.820	0.846	-3.2#	97	0.00
53 T	t-1,3-Dichloropropene	0.510	0.543	-6.5	98	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.582	-8.4	99	0.00
55 T	1,1,2-Trichloroethane	0.325	0.339	-4.3	100	0.00
56 T	Ethyl methacrylate	0.496	0.488	1.6	92	0.00
57 T	1,3-Dichloropropane	0.516	0.548	-6.2	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.150	0.184	-22.7	104	0.00
59 T	2-Hexanone	0.299	0.303	-1.3	91	0.00
60 T	Dibromochloromethane	0.370	0.404	-9.2	102	0.00
61 T	1,2-Dibromoethane	0.340	0.350	-2.9	95	0.00
62 S	4-Bromofluorobenzene	0.416	0.429	-3.1	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.408	0.412	-1.0	89	0.00
65 PM	Chlorobenzene	0.983	1.022	-4.0	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.401	-5.5	102	0.00
67 C	Ethyl Benzene	1.784	1.833	-2.7#	99	0.00
68 T	m/p-Xylenes	0.674	0.692	-2.7	98	0.00
69 T	o-Xylene	0.660	0.664	-0.6	97	0.00
70 T	Styrene	1.104	1.131	-2.4	96	0.00
71 P	Bromoform	0.304	0.326	-7.2	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.667	3.638	0.8	99	0.00
74 T	N-amyl acetate	1.423	1.308	8.1	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.063	1.011	4.9	95	0.00
76 T	1,2,3-Trichloropropane	0.841	0.827	1.7	95	0.00
77 T	Bromobenzene	0.929	0.924	0.5	99	0.00
78 T	n-propylbenzene	4.116	4.163	-1.1	100	0.00
79 T	2-Chlorotoluene	2.448	2.415	1.3	100	0.00
80 T	1,3,5-Trimethylbenzene	3.084	3.087	-0.1	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.382	2.6	95	0.00
82 T	4-Chlorotoluene	2.456	2.529	-3.0	102	0.00
83 T	tert-Butylbenzene	2.667	2.633	1.3	99	0.00
84 T	1,2,4-Trimethylbenzene	3.051	3.056	-0.2	100	0.00
85 T	sec-Butylbenzene	3.504	3.551	-1.3	100	0.00
86 T	p-Isopropyltoluene	3.243	3.298	-1.7	102	0.00
87 T	1,3-Dichlorobenzene	1.613	1.679	-4.1	103	0.00
88 T	1,4-Dichlorobenzene	1.609	1.673	-4.0	103	0.00
89 T	n-Butylbenzene	2.765	2.920	-5.6	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN013120\
Data File : VN059938.D
Acq On : 31 Jan 2020 10:54
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 03 01:57:27 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.607	0.614	-1.2	100	0.00
91 T	1,2-Dichlorobenzene	1.595	1.620	-1.6	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.207	19.5	89	0.00
93 T	1,2,4-Trichlorobenzene	0.958	1.003	-4.7	97	0.00
94 T	Hexachlorobutadiene	0.528	0.534	-1.1	100	0.00
95 T	Naphthalene	2.726	2.495	8.5	84	0.00
96 T	1,2,3-Trichlorobenzene	0.933	0.940	-0.8	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6