

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013120\
 Data File : VN059956.D
 Acq On : 31 Jan 2020 18:21
 Operator : JC/MD
 Sample : PB126487ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB126487ZHE#07

Quant Time: Feb 03 03:38:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	190269	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	295761	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	273070	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	124911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	107825	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
35) Dibromofluoromethane	7.57	113	89952	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
50) Toluene-d8	10.08	98	352530	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	12.40	95	125306	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
Target Compounds						
20) Methylene Chloride	4.52	84	6947	2.402	ug/l #	82
43) Isopropyl Acetate	8.14	43	40682	9.791	ug/l #	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN013120\
Data File : VN059956.D
Acq On : 31 Jan 2020 18:21
Operator : JC/MD
Sample : PB126487ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB126487ZHE#07

Quant Time: Feb 03 03:38:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21

Instrument :

MSVOA_N

ClientSampleId :


PB126487ZHE#07

Tot Ion:168 Resp: 190269

Ion Ratio Lower Upper




168 100

99 52.9 48.2 72.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#20

Methylene Chloride

Concen: 2.402 ug/l

RT: 4.52 min Scan# 855

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21

Tgt Ion: 84 Resp: 6947


Ion Ratio Lower Upper

84 100

49 99.4 106.6 160.0#

51 34.4 31.9 47.9

86 65.8 50.8 76.2






Abundance Ion 83.90 (83.60 to 84.60): VNO

Ion 48.90 (48.60 to 49.60): VNO

Ion 50.90 (50.60 to 51.60): VNO

Ion 85.90 (85.60 to 86.60): VNO









#33
 1,2-Dichloroethane-d4
 Concen: 51.030 ug/l
 RT: 8.00 min Scan# 1938
 Delta R.T. 0.00 min
 Lab File: VN059956.D
 Acq: 31 Jan 2020 18:21




Instrument :
 MSVOA_N
 ClientSampleId :
 PB126487ZHE#07


Tot Ion: 65 Resp: 107825
 Ion Ratio Lower Upper
 65 100
 67 52.5 0.0 103.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2113
 Delta R.T. 0.00 min
 Lab File: VN059956.D
 Acq: 31 Jan 2020 18:21

Tgt Ion: 114 Resp: 295761
 Ion Ratio Lower Upper
 114 100
 63 19.3 0.0 43.6
 88 15.5 0.0 31.4





#35

Dibromofluoromethane

Concen: 51.000 ug/l

RT: 7.57 min Scan# 1802

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21

Instrument :

MSVOA_N

ClientSampleId :

PB126487ZHE#07


Tot Ion:113 Resp: 89952

Ion Ratio Lower Upper

113 100

111 104.2 83.5 125.3




192 20.8 15.4 23.2



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#43

Isopropyl Acetate

Concen: 9.791 ug/l

RT: 8.14 min Scan# 1981

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21


Tgt Ion: 43 Resp: 40682

Ion Ratio Lower Upper

43 100

61 20.0 20.2 30.2#



87 14.4 9.5 14.3#




Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 61.00 (60.70 to 61.70): VNO

Ion 87.00 (86.70 to 87.70): VNO





#50

Toluene-d8

Concen: 51.587 ug/l

RT: 10.08 min Scan# 2583

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21

Instrument :

MSVOA_N

ClientSampleId :





PB126487ZHE#07

Tot Ion: 98 Resp: 352530

Ion Ratio Lower Upper

98 100

100 64.8 51.2 76.8



#62

4-Bromofluorobenzene

Concen: 50.864 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21




Tgt Ion: 95 Resp: 125306


Ion Ratio Lower Upper

95 100

174 88.0 0.0 151.4

176 84.8 0.0 146.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2994

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21

Instrument :

MSVOA_N

ClientSampleId :

PB126487ZHE#07


Tot Ion:117 Resp: 273070

Ion Ratio Lower Upper

117 100

82 52.8 45.9 68.9




119 32.3 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN059956.D

Acq: 31 Jan 2020 18:21


Tgt Ion:152 Resp: 124911

Ion Ratio Lower Upper

152 100

115 57.7 30.3 90.9

150 157.1 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

