

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070847.D
 Acq On : 31 Jan 2022 12:20
 Operator : JC/MD
 Sample : VN0131WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBL01

Quant Time: Jan 31 16:42:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	873889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1382281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1193072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	420205	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	585608	46.758	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.520%
35) Dibromofluoromethane	8.021	113	409121	47.715	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.420%
50) Toluene-d8	10.438	98	1685413	46.660	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.320%
62) 4-Bromofluorobenzene	12.728	95	592869	46.060	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
Data File : VN070847.D
Acq On : 31 Jan 2022 12:20
Operator : JC/MD
Sample : VN0131WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Quant Time: Jan 31 16:42:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

