

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070850.D
 Acq On : 31 Jan 2022 14:17
 Operator : JC/MD
 Sample : N1314-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-012722

Quant Time: Jan 31 16:43:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.078	168	790991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1261329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1089716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	441890	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	553562	48.832	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.660%
35) Dibromofluoromethane	8.016	113	364527	46.591	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.180%
50) Toluene-d8	10.438	98	1563035	47.421	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.840%
62) 4-Bromofluorobenzene	12.731	95	517784	44.084	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
Data File : VN070850.D
Acq On : 31 Jan 2022 14:17
Operator : JC/MD
Sample : N1314-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-01-012722

Quant Time: Jan 31 16:43:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

