

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070854.D
 Acq On : 31 Jan 2022 16:08
 Operator : JC/MD
 Sample : N1314-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-012722

Quant Time: Jan 31 16:45:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.078	168	806252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1304252	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1157284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	487602	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	560783	48.532	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.060%
35) Dibromofluoromethane	8.016	113	370668	45.816	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.640%
50) Toluene-d8	10.438	98	1632431	47.897	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.800%
62) 4-Bromofluorobenzene	12.731	95	564910	46.514	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
Data File : VN070854.D
Acq On : 31 Jan 2022 16:08
Operator : JC/MD
Sample : N1314-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-01-012722

Quant Time: Jan 31 16:45:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

