

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070856.D
 Acq On : 31 Jan 2022 16:59
 Operator : JC/MD
 Sample : N1284-01
 Misc : 3.47g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-1

Quant Time: Jan 31 17:16:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	821142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1321729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1229741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	538154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	560044	47.589	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.180%
35) Dibromofluoromethane	8.018	113	372379	45.419	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.840%
50) Toluene-d8	10.438	98	1659922	48.059	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.120%
62) 4-Bromofluorobenzene	12.731	95	590485	47.977	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.960%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.851	43	106277	6.616	ug/l	92
67) Ethyl Benzene	11.843	91	12112	0.245	ug/l #	84
68) m/p-Xylenes	11.948	106	4346	0.242	ug/l	80
84) 1,2,4-Trimethylbenzene	13.366	105	16796	0.424	ug/l	97
86) p-Isopropyltoluene	13.613	119	58583	1.531	ug/l	93
95) Naphthalene	15.504	128	97087	8.083	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
Data File : VN070856.D
Acq On : 31 Jan 2022 16:59
Operator : JC/MD
Sample : N1284-01
Misc : 3.47g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-1

Quant Time: Jan 31 17:16:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

