

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070858.D
 Acq On : 31 Jan 2022 17:49
 Operator : JC/MD
 Sample : N1319-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-2D

Quant Time: Feb 01 05:01:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	859390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1352260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1224234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	450835	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	576700	46.824	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.640%
35) Dibromofluoromethane	8.021	113	393238	46.881	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.760%
50) Toluene-d8	10.440	98	1705891	48.275	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.560%
62) 4-Bromofluorobenzene	12.731	95	560309	44.497	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.000%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	96952	11.341	ug/l	99
43) Isopropyl Acetate	8.560	43	98529	3.441	ug/l #	87
95) Naphthalene	15.507	128	30106	6.109	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
Data File : VN070858.D
Acq On : 31 Jan 2022 17:49
Operator : JC/MD
Sample : N1319-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-2D

Quant Time: Feb 01 05:01:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

