

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070859.D
 Acq On : 31 Jan 2022 18:15
 Operator : JC/MD
 Sample : N1313-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 17B-MISC-A

Quant Time: Feb 01 05:01:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	912264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1432027	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1289439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	471669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	611565	46.777	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.560%
35) Dibromofluoromethane	8.021	113	421920	47.498	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.000%
50) Toluene-d8	10.440	98	1794293	47.948	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.900%
62) 4-Bromofluorobenzene	12.731	95	584538	43.835	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.680%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	89656	9.880	ug/l	100
43) Isopropyl Acetate	8.560	43	99692	3.288	ug/l #	88
95) Naphthalene	15.504	128	37890	6.344	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
Data File : VN070859.D
Acq On : 31 Jan 2022 18:15
Operator : JC/MD
Sample : N1313-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
17B-MISC-A

Quant Time: Feb 01 05:01:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

