

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070860.D
 Acq On : 31 Jan 2022 18:40
 Operator : JC/MD
 Sample : N1287-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WASTE-CHARACTER

Quant Time: Feb 01 05:01:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	940106	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1484108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1334142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	493419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	631459	46.868	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.740%
35) Dibromofluoromethane	8.021	113	430257	46.737	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.480%
50) Toluene-d8	10.440	98	1853477	47.792	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.580%
62) 4-Bromofluorobenzene	12.731	95	612155	44.295	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.600%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	106583	11.398	ug/l	97
43) Isopropyl Acetate	8.560	43	103087	3.280	ug/l #	85
95) Naphthalene	15.504	128	11182	5.342	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
Data File : VN070860.D
Acq On : 31 Jan 2022 18:40
Operator : JC/MD
Sample : N1287-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WASTE-CHARACTER

Quant Time: Feb 01 05:01:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

