


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013122\
 Data File : VN070844.D
 Acq On : 31 Jan 2022 9:39
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Title : SW846 8260
 Last Update : Tue Jan 18 13:32:07 2022



AutoFind: Scans 4003, 4004, 4005; Background Corrected with Scan 3986

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	25099	PASS
75	95	30	60	55.3	67144	PASS
95	95	100	100	100.0	121355	PASS
96	95	5	9	6.6	8003	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.4	91496	PASS
175	174	5	9	8.0	7346	PASS
176	174	95	101	95.7	87531	PASS
177	176	5	9	6.8	5924	PASS