

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
 Data File : VN080842.D
 Acq On : 31 Jan 2024 15:56
 Operator : JC\MD
 Sample : P1276-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-36

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/01/2024
 Supervised By : Mahesh Dadoda 02/01/2024

Quant Time: Jan 31 23:18:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	92543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	222470	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	260207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	79319	49.544	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.080%		
35) Dibromofluoromethane	8.165	113	59559	43.759	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	87.520%		
50) Toluene-d8	10.570	98	245347	46.056	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.120%		
62) 4-Bromofluorobenzene	12.847	95	94944	51.145	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	102.280%		
Target Compounds						
16) Acetone	4.424	43	23829	41.040	ug/l	95
20) Methylene Chloride	5.277	84	4012m	2.632	ug/l	
37) Ethyl Acetate	7.565	43	2995	1.150	ug/l #	78
43) Isopropyl Acetate	8.688	43	114979	24.250	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
Data File : VN080842.D
Acq On : 31 Jan 2024 15:56
Operator : JC\MD
Sample : P1276-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-36

Quant Time: Jan 31 23:18:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024

