

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
 Data File : VN080847.D
 Acq On : 31 Jan 2024 17:55
 Operator : JC\MD
 Sample : P1277-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-41

Quant Time: Jan 31 23:20:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	88029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	209455	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	222993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	72328	47.494	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.980%
35) Dibromofluoromethane	8.165	113	55607	43.394	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.780%
50) Toluene-d8	10.571	98	229002	45.659	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.320%
62) 4-Bromofluorobenzene	12.853	95	88843	50.832	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.660%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	18849	34.128	ug/l #	88
20) Methylene Chloride	5.271	84	4888	3.371	ug/l #	78
43) Isopropyl Acetate	8.688	43	113651	25.459	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
Data File : VN080847.D
Acq On : 31 Jan 2024 17:55
Operator : JC\MD
Sample : P1277-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-41

Quant Time: Jan 31 23:20:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

