

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
 Data File : VN080848.D
 Acq On : 31 Jan 2024 18:19
 Operator : JC\MD
 Sample : P1284-09 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31913

Quant Time: Jan 31 23:20:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	85355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	194678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	213173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	86348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	73736	49.936	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.880%		
35) Dibromofluoromethane	8.165	113	56122	47.120	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.240%		
50) Toluene-d8	10.571	98	231020	49.557	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.120%		
62) 4-Bromofluorobenzene	12.853	95	88554	54.513	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.020%		
Target Compounds					Qvalue	
16) Acetone	4.424	43	11698	21.844	ug/l	96
25) 2-Butanone	7.488	43	4016	3.948	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
Data File : VN080848.D
Acq On : 31 Jan 2024 18:19
Operator : JC\MD
Sample : P1284-09 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31913

Quant Time: Jan 31 23:20:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

