

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
 Data File : VN080833.D
 Acq On : 31 Jan 2024 12:11
 Operator : JC\MD
 Sample : VN0131WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBL01

Quant Time: Jan 31 23:13:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	90070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	214537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	225598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	85367	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	73976	47.476	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.960%
35) Dibromofluoromethane	8.165	113	58851	44.837	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.680%
50) Toluene-d8	10.570	98	239719	46.663	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.320%
62) 4-Bromofluorobenzene	12.847	95	93076	51.992	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.980%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
Data File : VN080833.D
Acq On : 31 Jan 2024 12:11
Operator : JC\MD
Sample : VN0131WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Quant Time: Jan 31 23:13:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN080833.D
Acq: 31 Jan 2024 12:11




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01


Tgt Ion:168 Resp: 90070
Ion Ratio Lower Upper
168 100
99 69.1 59.9 89.9



#33
1,2-Dichloroethane-d4
Concen: 47.476 ug/l
RT: 8.582 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VN080833.D
Acq: 31 Jan 2024 12:11

Tgt Ion: 65 Resp: 73976
Ion Ratio Lower Upper
65 100
67 55.0 0.0 102.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN080833.D

Acq: 31 Jan 2024 12:11

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBL01





Tgt Ion:114 Resp: 214537

Ion Ratio Lower Upper

114 100

63 24.8 0.0 48.0

88 18.1 0.0 34.8



#35

Dibromofluoromethane

Concen: 44.837 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.006 min

Lab File: VN080833.D

Acq: 31 Jan 2024 12:11




Tgt Ion:113 Resp: 58851


Ion Ratio Lower Upper

113 100

111 102.0 82.2 123.4

192 14.8 12.5 18.7









#50
Toluene-d8
Concen: 46.663 ug/l
RT: 10.570 min Scan# 1465
Delta R.T. 0.006 min
Lab File: VN080833.D
Acq: 31 Jan 2024 12:11




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01


Tgt Ion: 98 Resp: 239719
Ion Ratio Lower Upper
98 100
100 63.7 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 51.992 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN080833.D
Acq: 31 Jan 2024 12:11

Tgt Ion: 95 Resp: 93076
Ion Ratio Lower Upper
95 100
174 62.0 0.0 120.4
176 59.5 0.0 121.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.870 min Scan# 117
Delta R.T. 0.006 min
Lab File: VN080833.D
Acq: 31 Jan 2024 12:11

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBL01

Tgt Ion:117 Resp: 225598
Ion Ratio Lower Upper
117 100
82 62.9 52.7 79.1
119 31.8 25.3 37.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN080833.D
Acq: 31 Jan 2024 12:11

Tgt Ion:152 Resp: 85367
Ion Ratio Lower Upper
152 100
115 68.4 32.3 96.8
150 161.3 0.0 339.8

