

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
 Data File : VN080834.D
 Acq On : 31 Jan 2024 12:35
 Operator : JC\MD
 Sample : VN0131WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/01/2024
 Supervised By : Mahesh Dadoda 02/01/2024

Quant Time: Jan 31 23:14:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	110119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	220965	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	190064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	99868	52.423	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.840%			
35) Dibromofluoromethane	8.165	113	71875	53.167	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.340%			
50) Toluene-d8	10.565	98	280135	52.944	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.880%			
62) 4-Bromofluorobenzene	12.853	95	94838	51.436	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.880%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.124	85	31804	19.445	ug/l	90
3) Chloromethane	2.359	50	52570	21.382	ug/l	96
4) Vinyl Chloride	2.512	62	38132	20.400	ug/l	91
5) Bromomethane	2.959	94	16375	21.902	ug/l	99
6) Chloroethane	3.124	64	23799	23.882	ug/l #	88
7) Trichlorofluoromethane	3.501	101	49762	20.570	ug/l	98
8) Diethyl Ether	3.953	74	23443	21.007	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	30295	21.476	ug/l	98
10) Methyl Iodide	4.595	142	23269	19.540	ug/l	92
11) Tert butyl alcohol	5.518	59	31862	102.602	ug/l #	82
12) 1,1-Dichloroethene	4.348	96	28138	18.204	ug/l	93
13) Acrolein	4.177	56	42999	109.801	ug/l	97
14) Allyl chloride	5.024	41	64143	20.994	ug/l	95
15) Acrylonitrile	5.718	53	97389	104.276	ug/l	98
16) Acetone	4.430	43	72591	105.067	ug/l	89
17) Carbon Disulfide	4.712	76	89790	18.659	ug/l	97
18) Methyl Acetate	5.024	43	52638	21.861	ug/l	97
19) Methyl tert-butyl Ether	5.794	73	108860	19.988	ug/l	96
20) Methylene Chloride	5.283	84	36033	19.868	ug/l	97
21) trans-1,2-Dichloroethene	5.777	96	32724	19.487	ug/l	98
22) Diisopropyl ether	6.671	45	144307	21.407	ug/l	97
23) Vinyl Acetate	6.606	43	534308	109.544	ug/l	100
24) 1,1-Dichloroethane	6.571	63	72084	21.085	ug/l	97
25) 2-Butanone	7.483	43	132402	100.892	ug/l	94
26) 2,2-Dichloropropane	7.488	77	56971	20.617	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	38966	20.045	ug/l	96
28) Bromochloromethane	7.806	49	34328	18.965	ug/l	92
29) Tetrahydrofuran	7.847	42	91951	103.550	ug/l	97
30) Chloroform	7.971	83	63652	20.041	ug/l	93
31) Cyclohexane	8.265	56	63951	19.577	ug/l	92
32) 1,1,1-Trichloroethane	8.171	97	53133	20.629	ug/l	99
36) 1,1-Dichloropropene	8.371	75	50587	20.133	ug/l	97
37) Ethyl Acetate	7.559	43	59276	22.916	ug/l	97
38) Carbon Tetrachloride	8.359	117	39455	19.489	ug/l	96
39) Methylcyclohexane	9.600	83	50055	20.042	ug/l	97
40) Benzene	8.612	78	154044	20.859	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
 Data File : VN080834.D
 Acq On : 31 Jan 2024 12:35
 Operator : JC\MD
 Sample : VN0131WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/01/2024
 Supervised By : Mahesh Dadoda 02/01/2024

Quant Time: Jan 31 23:14:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	32544	21.347	ug/l	93
42) 1,2-Dichloroethane	8.677	62	54696	22.734	ug/l	97
43) Isopropyl Acetate	8.688	43	97486	20.701	ug/l	98
44) Trichloroethene	9.353	130	31321	19.255	ug/l	91
45) 1,2-Dichloropropane	9.624	63	42487	21.259	ug/l	98
46) Dibromomethane	9.712	93	23910	20.182	ug/l	99
47) Bromodichloromethane	9.888	83	47722	19.840	ug/l #	95
48) Methyl methacrylate	9.677	41	44656	19.560	ug/l	97
49) 1,4-Dioxane	9.700	88	12045	437.991	ug/l	98
51) 4-Methyl-2-Pentanone	10.447	43	277950	105.579	ug/l	100
52) Toluene	10.635	92	89397	21.113	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	57463	20.507	ug/l	94
54) cis-1,3-Dichloropropene	10.312	75	62206	20.415	ug/l	96
55) 1,1,2-Trichloroethane	11.012	97	33904	20.835	ug/l	92
56) Ethyl methacrylate	10.871	69	56291	22.502	ug/l	98
57) 1,3-Dichloropropane	11.165	76	62439	20.388	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	141454	96.998	ug/l	100
59) 2-Hexanone	11.194	43	212422	103.877	ug/l	98
60) Dibromochloromethane	11.359	129	30945	20.444	ug/l	96
61) 1,2-Dibromoethane	11.476	107	31321	20.537	ug/l	93
64) Tetrachloroethene	11.106	164	24719	19.787	ug/l	96
65) Chlorobenzene	11.894	112	86624	20.883	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.965	131	27352	20.277	ug/l	96
67) Ethyl Benzene	11.965	91	160841	20.862	ug/l	93
68) m/p-Xylenes	12.071	106	117302	42.842	ug/l	99
69) o-Xylene	12.400	106	58690	21.572	ug/l	98
70) Styrene	12.412	104	93397	21.375	ug/l	99
71) Bromoform	12.582	173	18062	19.803	ug/l #	94
73) Isopropylbenzene	12.694	105	136721	20.713	ug/l	98
74) N-amyl acetate	12.494	43	84707	20.670	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.941	83	46132	19.869	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	39292m	18.074	ug/l	
77) Bromobenzene	12.982	156	27346	19.381	ug/l	92
78) n-propylbenzene	13.035	91	166375	21.099	ug/l	99
79) 2-Chlorotoluene	13.129	91	103750	20.701	ug/l	98
80) 1,3,5-Trimethylbenzene	13.176	105	110841	21.085	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.741	75	18821	22.304	ug/l	90
82) 4-Chlorotoluene	13.223	91	100104	20.171	ug/l	97
83) tert-Butylbenzene	13.441	119	90150	21.475	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	114219	21.445	ug/l	100
85) sec-Butylbenzene	13.618	105	126973	20.516	ug/l	98
86) p-Isopropyltoluene	13.729	119	100548	20.906	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	52090	20.640	ug/l	96
88) 1,4-Dichlorobenzene	13.818	146	49836	19.696	ug/l	97
89) n-Butylbenzene	14.059	91	96409	20.484	ug/l	97
90) Hexachloroethane	14.335	117	16353	19.432	ug/l	93
91) 1,2-Dichlorobenzene	14.106	146	50352	21.798	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	7705	17.754	ug/l	93
93) 1,2,4-Trichlorobenzene	15.394	180	22643	19.045	ug/l	93
94) Hexachlorobutadiene	15.506	225	10480	20.177	ug/l	90
95) Naphthalene	15.641	128	87164	19.160	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	22763	19.479	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013124\
Data File : VN080834.D
Acq On : 31 Jan 2024 12:35
Operator : JC\MD
Sample : VN0131WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01


Manual Integrations
APPROVED

Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024

Quant Time: Jan 31 23:14:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion:168 Resp: 110119
Ion Ratio Lower Upper
168 100
99 65.5 59.9 89.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#2
Dichlorodifluoromethane
Concen: 19.445 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion: 85 Resp: 31804
Ion Ratio Lower Upper
85 100
87 36.0 15.3 45.9





#3

Chloromethane

Concen: 21.382 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion: 50 Resp: 52570

Ion Ratio Lower Upper

50 100





52 33.6 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#4

Vinyl Chloride

Concen: 20.400 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN080834.D




Acq: 31 Jan 2024 12:35


Tgt Ion: 62 Resp: 38132

Ion Ratio Lower Upper

62 100


64 33.0 22.7 34.1





#5
Bromomethane
Concen: 21.902 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 94 Resp: 16379
Ion Ratio Lower Upper
94 100
96 87.1 69.2 103.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#6
Chloroethane
Concen: 23.882 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion: 64 Resp: 23799
Ion Ratio Lower Upper
64 100
66 26.2 26.4 39.6#









#7
Trichlorofluoromethane
Concen: 20.570 ug/l
RT: 3.501 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion:101 Resp: 49762
Ion Ratio Lower Upper
101 100
103 69.3 54.4 81.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#8
Diethyl Ether
Concen: 21.007 ug/l
RT: 3.953 min Scan# 340
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion: 74 Resp: 23443
Ion Ratio Lower Upper
74 100
45 106.1 53.3 159.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.476 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion:101 Resp: 3029
Ion Ratio Lower Upper
101 100
85 46.7 38.1 57.1
151 58.8 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#10
Methyl Iodide
Concen: 19.540 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.012 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion:142 Resp: 23269
Ion Ratio Lower Upper
142 100
127 40.6 37.0 55.4
141 13.4 12.5 18.7





#11
Tert butyl alcohol
Concen: 102.602 ug/l
RT: 5.518 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion: 59 Resp: 31862
Ion Ratio Lower Upper
59 100
57 4.0 8.6 12.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#12
1,1-Dichloroethene
Concen: 18.204 ug/l
RT: 4.348 min Scan# 407
Delta R.T. 0.012 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Tgt Ion: 96 Resp: 28138
Ion Ratio Lower Upper
96 100
61 184.0 138.2 207.4
98 63.2 49.2 73.8





#13
Acrolein
Concen: 109.801 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion: 56 Resp: 4299
Ion Ratio Lower Upper
56 100
55 71.2 55.3 82.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#14
Allyl chloride
Concen: 20.994 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Tgt Ion: 41 Resp: 64143
Ion Ratio Lower Upper
41 100
39 67.5 57.7 86.5
76 31.5 26.8 40.2





#15
Acrylonitrile
Concen: 104.276 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 53 Resp: 97389
Ion Ratio Lower Upper
53 100
52 82.8 64.8 97.2
51 32.9 26.9 40.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#16
Acetone
Concen: 105.067 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 43 Resp: 72591
Ion Ratio Lower Upper
43 100
58 25.2 24.8 37.2





#17
Carbon Disulfide
Concen: 18.659 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 76 Resp: 89790
Ion Ratio Lower Upper
76 100
78 9.0 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#18
Methyl Acetate
Concen: 21.861 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 43 Resp: 52638
Ion Ratio Lower Upper
43 100
74 20.1 17.3 25.9





#19
Methyl tert-butyl Ether
Concen: 19.988 ug/l
RT: 5.794 min Scan# 653
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 73 Resp: 108860
Ion Ratio Lower Upper
73 100
57 24.9 18.2 27.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#20
Methylene Chloride
Concen: 19.868 ug/l
RT: 5.283 min Scan# 566
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 84 Resp: 36033
Ion Ratio Lower Upper
84 100
49 139.0 108.1 162.1
51 46.2 32.9 49.3
86 60.5 48.6 73.0





#21
trans-1,2-Dichloroethene
Concen: 19.487 ug/l
RT: 5.777 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 96 Resp: 32724
Ion Ratio Lower Upper
96 100
61 138.0 112.8 169.2
98 63.0 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#22
Diisopropyl ether
Concen: 21.407 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 45 Resp: 144307
Ion Ratio Lower Upper
45 100
43 50.4 41.8 62.8
87 23.8 18.1 27.1
59 9.6 9.0 13.6





#23
 Vinyl Acetate
 Concen: 109.544 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBS01




Tgt Ion: 43 Resp: 534308
 Ion Ratio Lower Upper
 43 100
 86 8.7 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :John Carlone 02/01/2024
 Supervised By :Mahesh Dadoda 02/01/2024




#24
 1,1-Dichloroethane
 Concen: 21.085 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35



Tgt Ion: 63 Resp: 72084
 Ion Ratio Lower Upper
 63 100
 98 5.2 2.9 8.6
 100 5.5 1.9 5.7









#25
2-Butanone
Concen: 100.892 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion: 43 Resp: 132402
Ion Ratio Lower Upper
43 100
72 21.5 19.5 29.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#26
2,2-Dichloropropane
Concen: 20.617 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 77 Resp: 56971
Ion Ratio Lower Upper
77 100
97 21.9 11.5 34.5





#27
 cis-1,2-Dichloroethene
 Concen: 20.045 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBS01




Tgt Ion: 96 Resp: 38960
 Ion Ratio Lower Upper
 96 100
 61 168.3 0.0 325.4
 98 63.0 0.0 130.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 02/01/2024
 Supervised By :Mahesh Dadoda 02/01/2024




#28
 Bromochloromethane
 Concen: 18.965 ug/l
 RT: 7.806 min Scan# 995
 Delta R.T. -0.006 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35



Tgt Ion: 49 Resp: 34328
 Ion Ratio Lower Upper
 49 100
 129 1.0 0.0 1.6
 130 47.0 42.3 63.5





#29

Tetrahydrofuran

Concen: 103.550 ug/l

RT: 7.847 min Scan# 1000

Delta R.T. 0.012 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01





Tgt Ion:	42	Resp:	9195
Ion Ratio	Lower	Upper	
42	100		
72	37.2	31.1	46.7
71	33.4	28.3	42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#30

Chloroform

Concen: 20.041 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. 0.006 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35


Tgt Ion:	83	Resp:	63652
Ion Ratio	Lower	Upper	
83	100		
85	69.1	51.0	76.6





#31
Cyclohexane
Concen: 19.577 ug/l
RT: 8.265 min Scan# 1073
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 56 Resp: 6395
Ion Ratio Lower Upper
56 100
69 26.7 20.1 30.1
84 63.0 56.9 85.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#32
1,1,1-Trichloroethane
Concen: 20.629 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion: 97 Resp: 53133
Ion Ratio Lower Upper
97 100
99 67.1 53.4 80.0
61 53.7 43.8 65.8





#33

1,2-Dichloroethane-d4

Concen: 52.423 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion: 65 Resp: 9986

Ion Ratio Lower Upper

65 100





67 55.3 0.0 102.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35




Tgt Ion:114 Resp: 220965


Ion Ratio Lower Upper

114 100

63 24.9 0.0 48.0


88 16.1 0.0 34.8





#35
Dibromofluoromethane
Concen: 53.167 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 113 Resp: 71875
Ion Ratio Lower Upper
113 100
111 101.0 82.2 123.4
192 14.6 12.5 18.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#36
1,1-Dichloropropene
Concen: 20.133 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion: 75 Resp: 50587
Ion Ratio Lower Upper
75 100
110 31.0 16.0 48.0
77 29.4 25.8 38.6





#37

Ethyl Acetate

Concen: 22.916 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01





Tgt Ion:	43	Resp:	59270
Ion	Ratio	Lower	Upper
43	100		
61	11.6	10.6	15.8
70	9.3	8.0	12.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#38

Carbon Tetrachloride

Concen: 19.489 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. -0.006 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Tgt Ion:	117	Resp:	39455
Ion	Ratio	Lower	Upper
117	100		
119	89.7	73.4	110.2
121	25.6	24.4	36.6





#39

Methylcyclohexane

Concen: 20.042 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01





Tgt Ion: 83	Resp: 5005
Ion Ratio	Lower Upper
83	100
55	85.0 65.7 98.5
98	40.6 33.8 50.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#40

Benzene

Concen: 20.859 ug/l




RT: 8.612 min Scan# 1132


Delta R.T. 0.006 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35


Tgt Ion: 78	Resp: 154044
Ion Ratio	Lower Upper
78	100
77	22.5 20.4 30.6





#41
Methacrylonitrile
Concen: 21.347 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 41 Resp: 32544
Ion Ratio Lower Upper
41 100
39 52.9 48.4 72.6
67 64.8 56.8 85.2
52 30.4 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#42
1,2-Dichloroethane
Concen: 22.734 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion: 62 Resp: 54696
Ion Ratio Lower Upper
62 100
98 7.8 0.0 18.0





#43

Isopropyl Acetate

Concen: 20.701 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion: 43 Resp: 97486





Ion	Ratio	Lower	Upper
43	100		
61	24.2	19.9	29.9
87	10.4	9.0	13.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#44

Trichloroethene

Concen: 19.255 ug/l

RT: 9.353 min Scan# 1258




Delta R.T. 0.000 min


Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Tgt Ion:130 Resp: 31321

Ion	Ratio	Lower	Upper
130	100		
95	114.7	0.0	210.4





#45

1,2-Dichloropropane

Concen: 21.259 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion: 63 Resp: 4248

Ion Ratio Lower Upper

63 100





65 33.8 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#46

Dibromomethane

Concen: 20.182 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35




Tgt Ion: 93 Resp: 23910


Ion Ratio Lower Upper

93 100

95 83.0 68.1 102.1


174 71.1 57.2 85.8





#47
Bromodichloromethane
Concen: 19.840 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 83 Resp: 4772
Ion Ratio Lower Upper
83 100
85 61.6 52.3 78.5
127 8.2 4.6 7.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#48
Methyl methacrylate
Concen: 19.560 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion: 41 Resp: 44656
Ion Ratio Lower Upper
41 100
69 75.6 62.2 93.2
39 61.7 47.6 71.4





#49

1,4-Dioxane

Concen: 437.991 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion: 88 Resp: 1204

Ion Ratio Lower Upper

88 100

43 39.0 30.1 45.1





58 81.1 63.9 95.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#50

Toluene-d8

Concen: 52.944 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN080834.D




Acq: 31 Jan 2024 12:35


Tgt Ion: 98 Resp: 280135

Ion Ratio Lower Upper

98 100

100 66.1 51.4 77.0









#51
4-Methyl-2-Pentanone
Concen: 105.579 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion: 43 Resp: 277950
Ion Ratio Lower Upper
43 100
58 36.8 29.4 44.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#52
Toluene
Concen: 21.113 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 92 Resp: 89397
Ion Ratio Lower Upper
92 100
91 174.5 139.4 209.0





#53
 t-1,3-Dichloropropene
 Concen: 20.507 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBS01




Tgt Ion: 75 Resp: 5746
 Ion Ratio Lower Upper
 75 100
 77 28.4 25.5 38.3

Manual Integrations
 APPROVED



Reviewed By :John Carlone 02/01/2024
 Supervised By :Mahesh Dadoda 02/01/2024




#54
 cis-1,3-Dichloropropene
 Concen: 20.415 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.000 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35




Tgt Ion: 75 Resp: 62206
 Ion Ratio Lower Upper
 75 100
 77 31.9 23.3 34.9
 39 58.0 44.6 66.8





#55
1,1,2-Trichloroethane
Concen: 20.835 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion:	97	Resp:	33904
Ion Ratio	Lower	Upper	
97	100		
83	88.3	72.9	109.3
85	50.7	50.2	75.2
99	55.7	49.6	74.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#56
Ethyl methacrylate
Concen: 22.502 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35



Tgt Ion:	69	Resp:	56291
Ion Ratio	Lower	Upper	
69	100		
41	79.5	61.0	91.6
39	39.7	31.8	47.6





#57

1,3-Dichloropropane

Concen: 20.388 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion: 76 Resp: 62439

Ion Ratio Lower Upper

76 100





78 32.5 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 96.998 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN080834.D




Acq: 31 Jan 2024 12:35


Tgt Ion: 63 Resp: 141454

Ion Ratio Lower Upper

63 100


106 23.4 18.6 28.0





#59
2-Hexanone
Concen: 103.877 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 43 Resp: 21242
Ion Ratio Lower Upper
43 100
58 50.8 25.9 77.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#60
Dibromochloromethane
Concen: 20.444 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion:129 Resp: 30945
Ion Ratio Lower Upper
129 100
127 75.1 39.4 118.2





#61
1,2-Dibromoethane
Concen: 20.537 ug/l
RT: 11.476 min Scan# 1618
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion:107 Resp: 3132
Ion Ratio Lower Upper
107 100
109 98.6 73.4 110.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#62
4-Bromofluorobenzene
Concen: 51.436 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Tgt Ion: 95 Resp: 94838
Ion Ratio Lower Upper
95 100
174 63.1 0.0 120.4
176 56.9 0.0 121.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion:117 Resp: 190064
Ion Ratio Lower Upper
117 100
82 66.7 52.7 79.1
119 31.4 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#64
Tetrachloroethene
Concen: 19.787 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Tgt Ion:164 Resp: 24719
Ion Ratio Lower Upper
164 100
166 138.9 105.8 158.6
129 115.0 89.6 134.4
131 101.8 84.9 127.3





#65
Chlorobenzene
Concen: 20.883 ug/l
RT: 11.894 min Scan# 111
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion:112 Resp: 86624
Ion Ratio Lower Upper
112 100
114 33.5 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 20.277 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion:131 Resp: 27352
Ion Ratio Lower Upper
131 100
133 90.2 46.8 140.3
119 72.9 34.6 103.9









#67
Ethyl Benzene
Concen: 20.862 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion: 91 Resp: 16084
Ion Ratio Lower Upper
91 100
106 33.4 23.8 35.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#68
m/p-Xylenes
Concen: 42.842 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion:106 Resp: 117302
Ion Ratio Lower Upper
106 100
91 212.7 169.0 253.6









#69
o-Xylene
Concen: 21.572 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion:106 Resp: 58690
Ion Ratio Lower Upper
106 100
91 220.8 112.1 336.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#70
Styrene
Concen: 21.375 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion:104 Resp: 93397
Ion Ratio Lower Upper
104 100
78 55.4 43.8 65.8
103 56.4 44.2 66.2





#71

Bromoform

Concen: 19.803 ug/l

RT: 12.582 min Scan# 1806

Delta R.T. 0.006 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01





Tgt Ion:	173	Resp:	1806
Ion Ratio	Lower	Upper	
173	100		
175	49.8	23.1	69.2
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.794 min Scan# 2013


Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Tgt Ion:	152	Resp:	72079
Ion Ratio	Lower	Upper	
152	100		
115	67.7	32.3	96.8
150	172.5	0.0	339.8





#73

Isopropylbenzene

Concen: 20.713 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.006 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion:105 Resp: 13672

Ion Ratio Lower Upper

105 100





120 24.3 12.6 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#74

N-ethyl acetate

Concen: 20.670 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Tgt Ion: 43 Resp: 84707




Ion Ratio Lower Upper


43 100

70 35.3 30.1 45.1

55 24.4 17.0 25.4


61 23.3 17.5 26.3





#75
1,1,2,2-Tetrachloroethane
Concen: 19.869 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 83 Resp: 4613
Ion Ratio Lower Upper
83 100
131 7.5 4.1 12.3
85 65.1 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#76
1,2,3-Trichloropropane
Concen: 18.074 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 75 Resp: 39292
Ion Ratio Lower Upper
75 100
77 201.5 91.0 272.8





#77
Bromobenzene
Concen: 19.381 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion:156 Resp: 27340
Ion Ratio Lower Upper
156 100
77 289.5 138.1 414.3
158 102.1 46.5 139.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#78
n-propylbenzene
Concen: 21.099 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35



Tgt Ion: 91 Resp: 166375
Ion Ratio Lower Upper
91 100
120 21.2 10.3 30.9









#79
2-Chlorotoluene
Concen: 20.701 ug/l
RT: 13.129 min Scan# 1899
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion: 91 Resp: 103750
Ion Ratio Lower Upper
91 100
126 28.5 14.8 44.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#80
1,3,5-Trimethylbenzene
Concen: 21.085 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion:105 Resp: 110841
Ion Ratio Lower Upper
105 100
120 47.6 23.0 69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 22.304 ug/l
RT: 12.741 min Scan# 1833
Delta R.T. 0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 75 Resp: 1882
Ion Ratio Lower Upper
75 100
53 99.6 71.4 107.2
89 46.8 33.4 50.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#82
4-Chlorotoluene
Concen: 20.171 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 91 Resp: 100104
Ion Ratio Lower Upper
91 100
126 32.1 15.3 45.9





#83
tert-Butylbenzene
Concen: 21.475 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion:119 Resp: 90150
Ion Ratio Lower Upper
119 100
91 72.8 36.8 110.4
134 25.9 12.4 37.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#84
1,2,4-Trimethylbenzene
Concen: 21.445 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Tgt Ion:105 Resp: 114219
Ion Ratio Lower Upper
105 100
120 41.8 20.9 62.7





#85
 sec-Butylbenzene
 Concen: 20.516 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0131WBS01



Tgt Ion:105 Resp: 12697
 Ion Ratio Lower Upper
 105 100
 134 17.9 8.6 25.8




Manual Integrations
 APPROVED


Reviewed By :John Carlone 02/01/2024
 Supervised By :Mahesh Dadoda 02/01/2024



#86
 p-Isopropyltoluene
 Concen: 20.906 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN080834.D
 Acq: 31 Jan 2024 12:35

Tgt Ion:119 Resp: 100548
 Ion Ratio Lower Upper
 119 100
 134 24.6 12.4 37.2
 91 27.3 13.5 40.4





#87

1,3-Dichlorobenzene

Concen: 20.640 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01

Tgt Ion:146 Resp: 52090





Ion	Ratio	Lower	Upper
146	100		
111	43.5	23.9	71.7
148	63.8	32.6	97.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#88

1,4-Dichlorobenzene

Concen: 19.696 ug/l

RT: 13.818 min Scan# 2017




Delta R.T. 0.006 min


Lab File: VN080834.D

Acq: 31 Jan 2024 12:35

Tgt Ion:146 Resp: 49836


Ion	Ratio	Lower	Upper
146	100		
111	48.6	22.6	67.7
148	65.0	31.9	95.5





#89
n-Butylbenzene
Concen: 20.484 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01



Tgt Ion: 91 Resp: 96409
Ion Ratio Lower Upper
91 100
92 50.0 26.7 80.1
134 21.6 10.8 32.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#90
Hexachloroethane
Concen: 19.432 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Tgt Ion:117 Resp: 16353
Ion Ratio Lower Upper
117 100
201 54.9 30.0 90.0









#91
1,2-Dichlorobenzene
Concen: 21.798 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35

Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01

Tgt Ion:146 Resp: 50352
Ion Ratio Lower Upper
146 100
111 47.9 23.9 71.8
148 60.6 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.754 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35


Tgt Ion: 75 Resp: 7705
Ion Ratio Lower Upper
75 100
155 68.3 31.8 95.4
157 89.2 41.1 123.4





#93
1,2,4-Trichlorobenzene
Concen: 19.045 ug/l
RT: 15.394 min Scan# 21
Delta R.T. -0.006 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35




Instrument :
MSVOA_N
ClientSampleId :
VN0131WBS01




Tgt Ion:180 Resp: 2264
Ion Ratio Lower Upper
180 100
182 99.4 46.6 139.7
145 38.0 16.8 50.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 02/01/2024
Supervised By :Mahesh Dadoda 02/01/2024




#94
Hexachlorobutadiene
Concen: 20.177 ug/l
RT: 15.506 min Scan# 2304
Delta R.T. 0.000 min
Lab File: VN080834.D
Acq: 31 Jan 2024 12:35



Tgt Ion:225 Resp: 10480
Ion Ratio Lower Upper
225 100
223 59.7 33.3 99.8
227 57.3 33.1 99.2





#95

Naphthalene

Concen: 19.160 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.000 min

Lab File: VN080834.D


Acq: 31 Jan 2024 12:35

Instrument :

MSVOA_N

ClientSampleId :

VN0131WBS01






Tgt Ion:	128	Resp:	87164
Ion Ratio	Lower	Upper	
128	100		
127	12.7	10.0	15.0
129	10.0	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 02/01/2024

Supervised By :Mahesh Dadoda 02/01/2024



#96

1,2,3-Trichlorobenzene


Concen: 19.479 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.000 min

Lab File: VN080834.D

Acq: 31 Jan 2024 12:35



Tgt Ion:	180	Resp:	22763
Ion Ratio	Lower	Upper	
180	100		
182	93.3	46.5	139.3
145	36.8	17.8	53.4

