

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
 Data File : VN085602.D
 Acq On : 31 Jan 2025 15:15
 Operator : JC\MD
 Sample : PB166357ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166357ZHE#07

Quant Time: Jan 31 22:29:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175903	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	361116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	333777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	121720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	165798	58.391	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	116.780%	
35) Dibromofluoromethane	8.165	113	123541	49.313	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.620%	
50) Toluene-d8	10.565	98	457451	51.392	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.780%	
62) 4-Bromofluorobenzene	12.847	95	137139	45.040	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	90.080%	
Target Compounds						
16) Acetone	4.424	43	36794	40.105	ug/l	95
20) Methylene Chloride	5.283	84	9685	4.264	ug/l	86
43) Isopropyl Acetate	8.688	43	207687	36.522	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085602.D
Acq On : 31 Jan 2025 15:15
Operator : JC\MD
Sample : PB166357ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166357ZHE#07

Quant Time: Jan 31 22:29:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

