

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
 Data File : VN085603.D
 Acq On : 31 Jan 2025 15:39
 Operator : JC\MD
 Sample : PB166357ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166357ZHE#08

Quant Time: Jan 31 22:30:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	327744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	307543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	152340	59.700	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 119.400%			
35) Dibromofluoromethane	8.165	113	114569	50.388	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.780%			
50) Toluene-d8	10.565	98	425881	52.718	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.440%			
62) 4-Bromofluorobenzene	12.847	95	125681	45.479	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.960%			
Target Compounds						
16) Acetone	4.430	43	39522	47.935	ug/l	94
20) Methylene Chloride	5.265	84	7303	3.578	ug/l #	80
43) Isopropyl Acetate	8.688	43	213778	41.420	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085603.D
Acq On : 31 Jan 2025 15:39
Operator : JC\MD
Sample : PB166357ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166357ZHE#08

Quant Time: Jan 31 22:30:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

