

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
 Data File : VN085606.D
 Acq On : 31 Jan 2025 16:52
 Operator : JC\MD
 Sample : PB166357ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166357ZHE#11

Quant Time: Jan 31 22:31:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	343770	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	156269	57.455	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.920%	
35) Dibromofluoromethane	8.165	113	117254	49.165	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.320%	
50) Toluene-d8	10.565	98	436816	51.550	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.100%	
62) 4-Bromofluorobenzene	12.847	95	130030	44.860	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.720%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	38748	44.092	ug/l	98
20) Methylene Chloride	5.277	84	7764	3.569	ug/l	91
43) Isopropyl Acetate	8.688	43	228449	42.200	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085606.D
Acq On : 31 Jan 2025 16:52
Operator : JC\MD
Sample : PB166357ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166357ZHE#11

Quant Time: Jan 31 22:31:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

