

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
 Data File : VN085607.D
 Acq On : 31 Jan 2025 17:16
 Operator : JC\MD
 Sample : PB166357ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166357ZHE#12

Quant Time: Jan 31 22:32:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324359	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	149164	58.287	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	116.580%	
35) Dibromofluoromethane	8.165	113	112435	49.965	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.940%	
50) Toluene-d8	10.565	98	413718	51.746	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.500%	
62) 4-Bromofluorobenzene	12.847	95	122032	44.620	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.240%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	38673	46.770	ug/l	93
20) Methylene Chloride	5.283	84	7809	3.815	ug/l	96
43) Isopropyl Acetate	8.688	43	243046	47.583	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085607.D
Acq On : 31 Jan 2025 17:16
Operator : JC\MD
Sample : PB166357ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166357ZHE#12

Quant Time: Jan 31 22:32:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

