

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
 Data File : VN085608.D
 Acq On : 31 Jan 2025 17:40
 Operator : JC\MD
 Sample : PB166357ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166357ZHE#13

Quant Time: Jan 31 22:32:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110446	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	152211	60.152	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 120.300%			
35) Dibromofluoromethane	8.165	113	112986	50.663	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.320%			
50) Toluene-d8	10.565	98	416004	52.501	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.000%			
62) 4-Bromofluorobenzene	12.847	95	125869	46.438	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.880%			
Target Compounds						
16) Acetone	4.436	43	39464	48.268	ug/l	96
20) Methylene Chloride	5.283	84	8169	4.036	ug/l #	86
43) Isopropyl Acetate	8.688	43	243799	48.160	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085608.D
Acq On : 31 Jan 2025 17:40
Operator : JC\MD
Sample : PB166357ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166357ZHE#13

Quant Time: Jan 31 22:32:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

