

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
 Data File : VN085609.D
 Acq On : 31 Jan 2025 18:04
 Operator : JC\MD
 Sample : PB166357ZHE#14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166357ZHE#14

Quant Time: Jan 31 22:33:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	327919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	311174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	152218	59.810	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	119.620%		
35) Dibromofluoromethane	8.165	113	112569	49.482	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.960%		
50) Toluene-d8	10.565	98	417308	51.629	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.260%		
62) 4-Bromofluorobenzene	12.847	95	125609	45.429	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	90.860%		
Target Compounds						
16) Acetone	4.430	43	38873	47.272	ug/l	97
20) Methylene Chloride	5.271	84	7922	3.891	ug/l	93
43) Isopropyl Acetate	8.688	43	234200	45.353	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085609.D
Acq On : 31 Jan 2025 18:04
Operator : JC\MD
Sample : PB166357ZHE#14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166357ZHE#14

Quant Time: Jan 31 22:33:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

