

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
 Data File : VN085613.D
 Acq On : 31 Jan 2025 19:41
 Operator : JC\MD
 Sample : PB166357ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166357ZHE#01

Quant Time: Jan 31 22:34:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	344120	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	324187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160166	59.862	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	119.720%	
35) Dibromofluoromethane	8.165	113	118851	49.784	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.560%	
50) Toluene-d8	10.565	98	444373	52.389	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	104.780%	
62) 4-Bromofluorobenzene	12.847	95	130963	45.136	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	90.280%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	33178	38.378	ug/l	97
20) Methylene Chloride	5.265	84	3322	1.552	ug/l #	78
30) Chloroform	7.965	83	6392	1.583	ug/l	99
43) Isopropyl Acetate	8.688	43	261992	48.346	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013125\
Data File : VN085613.D
Acq On : 31 Jan 2025 19:41
Operator : JC\MD
Sample : PB166357ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166357ZHE#01

Quant Time: Jan 31 22:34:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

