

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020119\
 Data File : VN053581.D
 Acq On : 1 Feb 2019 17:22
 Operator : JC/SP
 Sample : K1011-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-013119-A

Manual Integrations
 APPROVED

MMDadoda
 2/4/2019 9:49:33 AM

Quant Time: Feb 01 23:43:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	788910	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1237523	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1142933	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	489674	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	397468	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
35) Dibromofluoromethane	7.59	113	391449	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
50) Toluene-d8	10.09	98	1497752	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	12.40	95	506092	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
Target Compounds						
16) Acetone	3.82	43	23715	11.682	ug/l	95
20) Methylene Chloride	4.55	84	2216281	260.224	ug/l	99
25) 2-Butanone	6.84	43	9434m	3.258	ug/l	
43) Isopropyl Acetate	8.17	43	31136	2.588	ug/l #	84
52) Toluene	10.16	92	57015	2.793	ug/l	98
95) Naphthalene	15.13	128	45096	3.760	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020119\
Data File : VN053581.D
Acq On : 1 Feb 2019 17:22
Operator : JC/SP
Sample : K1011-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-013119-A

Manual Integrations
APPROVED

MMDadoda
2/4/2019 9:49:33 AM

Quant Time: Feb 01 23:43:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

