

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020119\
 Data File : VN053573.D
 Acq On : 1 Feb 2019 13:27
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 01 22:59:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.515	0.554	-7.6	102	0.00
3 P	Chloromethane	0.594	0.539	9.3	94	0.00
4 C	Vinyl Chloride	0.576	0.551	4.3#	98	0.00
5 T	Bromomethane	0.377	0.346	8.2	97	0.00
6 T	Chloroethane	0.329	0.308	6.4	98	0.00
7 T	Trichlorofluoromethane	0.781	0.752	3.7	100	0.00
8 T	Diethyl Ether	0.285	0.266	6.7	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.480	0.6	105	0.00
10 T	Methyl Iodide	0.759	0.693	8.7	95	0.00
11 T	Tert butyl alcohol	0.030	0.029	3.3	104	0.00
12 CM	1,1-Dichloroethene	0.463	0.438	5.4#	98	0.00
13 T	Acrolein	0.016	0.010	37.5#	77	0.00
14 T	Allyl chloride	0.737	0.696	5.6	97	0.00
15 T	Acrylonitrile	0.160	0.155	3.1	99	0.00
16 T	Acetone	0.129	0.132	-2.3	113	0.00
17 T	Carbon Disulfide	1.442	1.336	7.4	97	0.00
18 T	Methyl Acetate	0.381	0.347	8.9	99	0.00
19 T	Methyl tert-butyl Ether	1.174	1.134	3.4	98	0.00
20 T	Methylene Chloride	0.540	0.486	10.0	98	0.00
21 T	trans-1,2-Dichloroethene	0.500	0.473	5.4	99	0.00
22 T	Diisopropyl ether	1.366	1.352	1.0	100	0.00
23 T	Vinyl Acetate	0.945	0.940	0.5	98	0.00
24 P	1,1-Dichloroethane	0.911	0.847	7.0	98	0.00
25 T	2-Butanone	0.183	0.183	0.0	103	0.00
26 T	2,2-Dichloropropane	0.696	0.651	6.5	98	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.526	5.2	97	0.00
28 T	Bromochloromethane	0.395	0.385	2.5	102	0.00
29 T	Tetrahydrofuran	0.118	0.116	1.7	99	0.00
30 C	Chloroform	0.903	0.843	6.6#	100	0.00
31 T	Cyclohexane	0.949	0.804	15.3	100	0.00
32 T	1,1,1-Trichloroethane	0.752	0.707	6.0	98	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.483	3.8	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.316	0.315	0.3	102	0.00
36 T	1,1-Dichloropropene	0.451	0.447	0.9	101	0.00
37 T	Ethyl Acetate	0.252	0.256	-1.6	96	0.00
38 T	Carbon Tetrachloride	0.454	0.434	4.4	98	0.00
39 T	Methylcyclohexane	0.535	0.558	-4.3	103	0.00
40 TM	Benzene	1.390	1.343	3.4	99	0.00
41 T	Methacrylonitrile	0.141	0.147	-4.3	107	0.00
42 TM	1,2-Dichloroethane	0.413	0.397	3.9	100	0.00
43 T	Isopropyl Acetate	0.486	0.451	7.2	97	0.00
44 TM	Trichloroethene	0.396	0.383	3.3	103	0.00
45 C	1,2-Dichloropropane	0.356	0.346	2.8#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020119\
 Data File : VN053573.D
 Acq On : 1 Feb 2019 13:27
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 01 22:59:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 01 09:11:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.202	5.6	98	0.00
47 T	Bromodichloromethane	0.446	0.437	2.0	101	0.00
48 T	Methyl methacrylate	0.232	0.220	5.2	97	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	102	0.00
50 S	Toluene-d8	1.186	1.178	0.7	101	0.00
51 T	4-Methyl-2-Pentanone	0.239	0.242	-1.3	99	0.00
52 CM	Toluene	0.825	0.820	0.6#	100	0.00
53 T	t-1,3-Dichloropropene	0.446	0.436	2.2	98	0.00
54 T	cis-1,3-Dichloropropene	0.529	0.517	2.3	98	0.00
55 T	1,1,2-Trichloroethane	0.297	0.291	2.0	99	0.00
56 T	Ethyl methacrylate	0.352	0.354	-0.6	97	0.00
57 T	1,3-Dichloropropane	0.508	0.493	3.0	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.183	6.6	88	0.00
59 T	2-Hexanone	0.162	0.163	-0.6	100	0.00
60 T	Dibromochloromethane	0.335	0.332	0.9	99	0.00
61 T	1,2-Dibromoethane	0.294	0.281	4.4	98	0.00
62 S	4-Bromofluorobenzene	0.396	0.396	0.0	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.429	0.416	3.0	100	0.00
65 PM	Chlorobenzene	1.062	1.021	3.9	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.361	2.7	99	0.00
67 C	Ethyl Benzene	1.745	1.711	1.9#	99	0.00
68 T	m/p-Xylenes	0.666	0.665	0.2	99	0.00
69 T	o-Xylene	0.643	0.636	1.1	98	0.00
70 T	Styrene	1.043	1.053	-1.0	99	0.00
71 P	Bromoform	0.254	0.255	-0.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.515	3.384	3.7	99	0.00
74 T	N-amyl acetate	0.845	0.788	6.7	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.841	0.768	8.7	97	0.00
76 T	1,2,3-Trichloropropane	0.642	0.632	1.6	103	0.00
77 T	Bromobenzene	0.947	0.906	4.3	99	0.00
78 T	n-propylbenzene	4.063	4.044	0.5	101	0.00
79 T	2-Chlorotoluene	2.411	2.341	2.9	101	0.00
80 T	1,3,5-Trimethylbenzene	2.915	2.936	-0.7	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.228	0.215	5.7	94	0.00
82 T	4-Chlorotoluene	2.474	2.408	2.7	100	0.00
83 T	tert-Butylbenzene	2.576	2.518	2.3	100	0.00
84 T	1,2,4-Trimethylbenzene	2.970	2.992	-0.7	101	0.00
85 T	sec-Butylbenzene	3.506	3.490	0.5	101	0.00
86 T	p-Isopropyltoluene	3.066	3.121	-1.8	102	0.00
87 T	1,3-Dichlorobenzene	1.703	1.633	4.1	101	0.00
88 T	1,4-Dichlorobenzene	1.685	1.597	5.2	99	0.00
89 T	n-Butylbenzene	2.654	2.729	-2.8	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020119\
Data File : VN053573.D
Acq On : 1 Feb 2019 13:27
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 01 22:59:57 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 01 09:11:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.506	0.486	4.0	100	0.00
91 T	1,2-Dichlorobenzene	1.630	1.536	5.8	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.114	5.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.935	0.979	-4.7	100	0.00
94 T	Hexachlorobutadiene	0.582	0.587	-0.9	103	0.00
95 T	Naphthalene	1.905	2.020	-6.0	95	0.00
96 T	1,2,3-Trichlorobenzene	0.868	0.912	-5.1	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6