

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020122\  
 Data File : VN070868.D  
 Acq On : 1 Feb 2022 12:57  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Feb 02 00:33:21 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011822W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 18 13:32:07 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0   | 90    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.578 | 0.504 | 12.8  | 81    | 0.00     |
| 3 P   | Chloromethane               | 0.845 | 0.704 | 16.7  | 79    | 0.00     |
| 4 C   | Vinyl Chloride              | 0.803 | 0.665 | 17.2# | 76    | 0.00     |
| 5 T   | Bromomethane                | 0.418 | 0.358 | 14.4  | 78    | -0.02    |
| 6 T   | Chloroethane                | 0.501 | 0.400 | 20.2  | 76    | -0.01    |
| 7 T   | Trichlorofluoromethane      | 0.918 | 0.861 | 6.2   | 87    | 0.00     |
| 8 T   | Diethyl Ether               | 0.401 | 0.381 | 5.0   | 89    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.555 | 0.528 | 4.9   | 91    | 0.00     |
| 10 T  | Methyl Iodide               | 0.715 | 0.703 | 1.7   | 87    | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.129 | 0.112 | 13.2  | 81    | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.553 | 0.503 | 9.0#  | 88    | 0.00     |
| 13 T  | Acrolein                    | 0.185 | 0.222 | -20.0 | 117   | 0.00     |
| 14 T  | Allyl chloride              | 1.205 | 1.169 | 3.0   | 86    | 0.00     |
| 15 T  | Acrylonitrile               | 0.411 | 0.381 | 7.3   | 82    | 0.00     |
| 16 T  | Acetone                     | 0.497 | 0.450 | 9.5   | 81    | 0.00     |
| 17 T  | Carbon Disulfide            | 1.502 | 1.322 | 12.0  | 80    | 0.00     |
| 18 T  | Methyl Acetate              | 0.978 | 0.901 | 7.9   | 86    | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 1.976 | 1.977 | -0.1  | 89    | 0.00     |
| 20 T  | Methylene Chloride          | 0.655 | 0.615 | 6.1   | 87    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.557 | 0.536 | 3.8   | 86    | 0.00     |
| 22 T  | Diisopropyl ether           | 2.389 | 2.353 | 1.5   | 88    | 0.00     |
| 23 T  | Vinyl Acetate               | 2.025 | 2.018 | 0.3   | 87    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 1.195 | 1.157 | 3.2   | 88    | 0.00     |
| 25 T  | 2-Butanone                  | 0.626 | 0.579 | 7.5   | 81    | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 0.916 | 0.938 | -2.4  | 93    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.661 | 0.650 | 1.7   | 89    | 0.00     |
| 28 T  | Bromochloromethane          | 0.536 | 0.509 | 5.0   | 89    | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.383 | 0.341 | 11.0  | 78    | 0.00     |
| 30 C  | Chloroform                  | 1.342 | 1.106 | 17.6# | 88    | 0.00     |
| 31 T  | Cyclohexane                 | 1.262 | 1.109 | 12.1  | 83    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 0.927 | 0.929 | -0.2  | 90    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.717 | 0.695 | 3.1   | 88    | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0   | 88    | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.310 | 0.332 | -7.1  | 93    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.522 | 0.524 | -0.4  | 88    | 0.00     |
| 37 T  | Ethyl Acetate               | 0.693 | 0.657 | 5.2   | 81    | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.464 | 0.487 | -5.0  | 91    | 0.00     |
| 39 T  | Methylcyclohexane           | 0.645 | 0.659 | -2.2  | 87    | 0.00     |
| 40 TM | Benzene                     | 1.618 | 1.607 | 0.7   | 87    | 0.00     |
| 41 T  | Methacrylonitrile           | 0.353 | 0.349 | 1.1   | 87    | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.561 | 0.563 | -0.4  | 88    | 0.00     |
| 43 T  | Isopropyl Acetate           | 1.059 | 1.064 | -0.5  | 86    | 0.00     |
| 44 TM | Trichloroethene             | 0.367 | 0.376 | -2.5  | 89    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.440 | 0.448 | -1.8# | 89    | 0.00     |
| 46 T  | Dibromomethane              | 0.256 | 0.263 | -2.7  | 88    | 0.00     |
| 47 T  | Bromodichloromethane        | 0.523 | 0.545 | -4.2  | 91    | 0.00     |
| 48 T  | Methyl methacrylate         | 0.504 | 0.483 | 4.2   | 79    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020122\  
 Data File : VN070868.D  
 Acq On : 1 Feb 2022 12:57  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Feb 02 00:33:21 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011822W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 18 13:32:07 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 0.008 | 0.007 | 12.5  | 77    | 0.00     |
| 50 S  | Toluene-d8                  | 1.307 | 1.340 | -2.5  | 90    | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.676 | 0.683 | -1.0  | 84    | 0.00     |
| 52 CM | Toluene                     | 0.968 | 0.990 | -2.3# | 88    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.561 | 0.619 | -10.3 | 92    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.627 | 0.682 | -8.8  | 91    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.380 | 0.389 | -2.4  | 90    | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.639 | 0.686 | -7.4  | 89    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.673 | 0.693 | -3.0  | 89    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.151 | 0.158 | -4.6  | 83    | 0.00     |
| 59 T  | 2-Hexanone                  | 0.491 | 0.490 | 0.2   | 80    | 0.00     |
| 60 T  | Dibromochloromethane        | 0.355 | 0.392 | -10.4 | 92    | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.382 | 0.394 | -3.1  | 89    | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.466 | 0.503 | -7.9  | 93    | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0   | 88    | 0.00     |
| 64 T  | Tetrachloroethene           | 0.367 | 0.355 | 3.3   | 87    | 0.00     |
| 65 PM | Chlorobenzene               | 1.083 | 1.116 | -3.0  | 91    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.373 | 0.397 | -6.4  | 92    | 0.00     |
| 67 C  | Ethyl Benzene               | 2.008 | 2.105 | -4.8# | 90    | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.731 | 0.774 | -5.9  | 90    | 0.00     |
| 69 T  | o-Xylene                    | 0.725 | 0.772 | -6.5  | 90    | 0.00     |
| 70 T  | Styrene                     | 1.160 | 1.278 | -10.2 | 91    | 0.00     |
| 71 P  | Bromoform                   | 0.282 | 0.311 | -10.3 | 93    | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 85    | 0.00     |
| 73 T  | Isopropylbenzene            | 4.730 | 4.877 | -3.1  | 92    | 0.00     |
| 74 T  | N-amyl acetate              | 1.934 | 1.898 | 1.9   | 81    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 1.598 | 1.453 | 9.1   | 82    | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 1.326 | 1.318 | 0.6   | 93    | 0.00     |
| 77 T  | Bromobenzene                | 1.061 | 1.088 | -2.5  | 90    | 0.00     |
| 78 T  | n-propylbenzene             | 5.290 | 5.730 | -8.3  | 92    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 3.346 | 3.417 | -2.1  | 90    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 3.825 | 4.011 | -4.9  | 91    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.445 | 0.448 | -0.7  | 85    | 0.00     |
| 82 T  | 4-Chlorotoluene             | 3.127 | 3.308 | -5.8  | 90    | 0.00     |
| 83 T  | tert-Butylbenzene           | 3.330 | 3.476 | -4.4  | 91    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 3.677 | 3.934 | -7.0  | 91    | 0.00     |
| 85 T  | sec-Butylbenzene            | 4.527 | 4.936 | -9.0  | 92    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.556 | 3.938 | -10.7 | 91    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.764 | 1.860 | -5.4  | 89    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 1.756 | 1.821 | -3.7  | 90    | 0.00     |
| 89 T  | n-Butylbenzene              | 2.934 | 3.376 | -15.1 | 95    | 0.00     |
| 90 T  | Hexachloroethane            | 0.622 | 0.675 | -8.5  | 92    | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 1.743 | 1.775 | -1.8  | 89    | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 0.271 | 0.259 | 4.4   | 82    | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 0.774 | 0.927 | -19.8 | 93    | 0.00     |
| 94 T  | Hexachlorobutadiene         | 0.516 | 0.568 | -10.1 | 94    | 0.00     |
| 95 T  | Naphthalene                 | 2.001 | 2.390 | -19.4 | 89    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020122\  
Data File : VN070868.D  
Acq On : 1 Feb 2022 12:57  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
LabSampleId :  
VSTDCCC050

Quant Time: Feb 02 00:33:21 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N011822W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jan 18 13:32:07 2022  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|------------------------|-------|-------|-------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 0.738 | 0.870 | -17.9 | 90    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6