

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020123\
 Data File : VN076496.D
 Acq On : 01 Feb 2023 14:02
 Operator : JC\MD
 Sample : 01383-03
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2876

Quant Time: Feb 01 23:27:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-5.55
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-6.76
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-10.06
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-12.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	78 - 117	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	8.594	98	119	0.000	ug/l	-0.05
Spiked Amount	50.000	Range	92 - 112	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	83 - 123	Recovery	=	0.000%#


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020123\
Data File : VN076496.D
Acq On : 01 Feb 2023 14:02
Operator : JC\MD
Sample : 01383-03
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2876

Quant Time: Feb 01 23:27:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 0.000 ug/l
Expected RT: 5.55 min

Lab File: VN076496.D
Acq: 01 Feb 2023 14:02

Instrument :
MSVOA_N
ClientSampleId :
2876


Tgt Ion: 168
Sig Exp Ratio
168 100
99 49.5




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 5.96 min

Lab File: VN076496.D
Acq: 01 Feb 2023 14:02

Tgt Ion: 65
Sig Exp Ratio
65 100
67 52.4







#34
 1,4-Difluorobenzene
 Concen: 0.000 ug/l
 Expected RT: 6.76 min
 Lab File: VN076496.D
 Acq: 01 Feb 2023 14:02


Instrument :
 MSVOA_N
 ClientSampleId :
 2876


Tgt Ion: 114	
Sig	Exp Ratio
114	100
63	18.7
88	15.3



#35
 Dibromofluoromethane
 Concen: 0.000 ug/l
 Expected RT: 5.39 min
 Lab File: VN076496.D
 Acq: 01 Feb 2023 14:02

Tgt Ion: 113	
Sig	Exp Ratio
113	100
111	103.8
192	23.7









#50
Toluene-d8
Concen: 0.000 ug/l
RT: 8.594 min Scan# 1112
Delta R.T. -0.053 min
Lab File: VN076496.D
Acq: 01 Feb 2023 14:02

Instrument :
MSVOA_N
ClientSampleId :
2876


Tgt Ion: 98 Resp: 119
Ion Ratio Lower Upper
98 100
100 699.2 53.6 80.4#




#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min

Lab File: VN076496.D
Acq: 01 Feb 2023 14:02

Tgt Ion: 95
Sig Exp Ratio
95 100
174 92.2
176 89.4







#63
 Chlorobenzene-d5
 Concen: 0.000 ug/l
 Expected RT: 10.06 min
 Lab File: VN076496.D
 Acq: 01 Feb 2023 14:02

Instrument :
 MSVOA_N
 ClientSampleId :
 2876

Tgt Ion: 117

Sig	Exp Ratio
117	100
82	50.7
119	32.0



#72
 1,4-Dichlorobenzene-d4
 Concen: 0.000 ug/l
 Expected RT: 12.02 min
 Lab File: VN076496.D
 Acq: 01 Feb 2023 14:02

Tgt Ion: 152

Sig	Exp Ratio
152	100
115	77.4
150	174.1

