

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020123\
 Data File : VN076490.D
 Acq On : 01 Feb 2023 11:17
 Operator : JC\MD
 Sample : VN0201WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0201WBL01

Quant Time: Feb 01 23:47:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	493052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	792349	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	706960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	279367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	283300	51.317	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.640%	
35) Dibromofluoromethane	8.171	113	242647	48.956	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.920%	
50) Toluene-d8	10.571	98	916216	50.053	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	12.853	95	301879	50.518	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.040%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020123\
Data File : VN076490.D
Acq On : 01 Feb 2023 11:17
Operator : JC\MD
Sample : VN0201WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0201WBL01

Quant Time: Feb 01 23:47:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

