

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
 Data File : VN080859.D
 Acq On : 01 Feb 2024 10:37
 Operator : JC\MD
 Sample : VN0201WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0201WBL01

Quant Time: Feb 02 03:31:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	86995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	210734	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	220769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89086	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	78469	52.139	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.280%
35) Dibromofluoromethane	8.165	113	59081	45.825	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.640%
50) Toluene-d8	10.565	98	243316	48.218	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.440%
62) 4-Bromofluorobenzene	12.847	95	93254	53.032	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
Data File : VN080859.D
Acq On : 01 Feb 2024 10:37
Operator : JC\MD
Sample : VN0201WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0201WBL01

Quant Time: Feb 02 03:31:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

