

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
 Data File : VN080862.D
 Acq On : 01 Feb 2024 12:02
 Operator : JC\MD
 Sample : PB158744TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB158744TB

Quant Time: Feb 02 03:36:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	76713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	189358	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	205067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	69623	52.462	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.920%	
35) Dibromofluoromethane	8.171	113	55220	47.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.340%	
50) Toluene-d8	10.571	98	214510	47.308	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.620%	
62) 4-Bromofluorobenzene	12.853	95	98358	62.249	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 124.500%#	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	18152	37.714	ug/l	98
20) Methylene Chloride	5.283	84	5498	4.352	ug/l #	72
37) Ethyl Acetate	7.559	43	5479	2.472	ug/l #	92
43) Isopropyl Acetate	8.683	43	128615	31.869	ug/l #	91
51) 4-Methyl-2-Pentanone	10.453	43	11512	5.103	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
Data File : VN080862.D
Acq On : 01 Feb 2024 12:02
Operator : JC\MD
Sample : PB158744TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB158744TB

Quant Time: Feb 02 03:36:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

