

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
 Data File : VN080867.D
 Acq On : 01 Feb 2024 14:01
 Operator : JC\MD
 Sample : P1311-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Feb 02 03:38:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/02/2024
 Supervised By : Semsettin Yesilyurt 02/02/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	84374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	209225	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	234496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	71030	48.662	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.320%		
35) Dibromofluoromethane	8.165	113	54885	42.877	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	85.760%		
50) Toluene-d8	10.571	98	239261	47.757	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.520%		
62) 4-Bromofluorobenzene	12.847	95	94628	54.201	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	108.400%		
Target Compounds						
16) Acetone	4.424	43	19732	37.274	ug/l	89
20) Methylene Chloride	5.277	84	6173m	4.442	ug/l	
37) Ethyl Acetate	7.559	43	14193	5.795	ug/l	96
43) Isopropyl Acetate	8.688	43	130753	29.323	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
Data File : VN080867.D
Acq On : 01 Feb 2024 14:01
Operator : JC\MD
Sample : P1311-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Feb 02 03:38:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/02/2024
Supervised By :Semsettin Yesilyurt 02/02/2024

