

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
 Data File : VN080869.D
 Acq On : 01 Feb 2024 14:49
 Operator : JC\MD
 Sample : P1311-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Feb 02 03:39:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 02/02/2024
 Supervised By : Semsettin Yesilyurt 02/02/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	83913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	199241	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	223331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	72085	49.656	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.320%		
35) Dibromofluoromethane	8.165	113	57413	47.100	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.200%		
50) Toluene-d8	10.571	98	228768	47.950	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.900%		
62) 4-Bromofluorobenzene	12.853	95	94847	57.049	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	114.100%		
Target Compounds						
16) Acetone	4.436	43	20507	38.951	ug/l	90
18) Methyl Acetate	5.018	43	3445m	1.878	ug/l	
20) Methylene Chloride	5.271	84	5741m	4.154	ug/l	
37) Ethyl Acetate	7.565	43	14678	6.293	ug/l #	91
43) Isopropyl Acetate	8.688	43	128374	30.232	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020124\
Data File : VN080869.D
Acq On : 01 Feb 2024 14:49
Operator : JC\MD
Sample : P1311-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Feb 02 03:39:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/02/2024
Supervised By :Semsettin Yesilyurt 02/02/2024

