

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070893.D
 Acq On : 2 Feb 2022 14:37
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Feb 03 02:42:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	831747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1383536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1307168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	569003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	1034134	91.853	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	183.700%#
35) Dibromofluoromethane	8.019	113	766605	92.494	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	184.980%#
50) Toluene-d8	10.440	98	3140469	92.285	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	184.560%#
62) 4-Bromofluorobenzene	12.731	95	1207180	98.450	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	196.900%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.070	85	885515	95.952	ug/l	100
3) Chloromethane	2.301	50	1149685	98.632	ug/l	99
4) Vinyl Chloride	2.443	62	919140	94.390	ug/l	99
5) Bromomethane	2.818	94	508855	102.168	ug/l	99
6) Chloroethane	2.995	64	421722	95.493	ug/l	98
7) Trichlorofluoromethane	3.363	101	1022303	94.010	ug/l	97
8) Diethyl Ether	3.819	74	604523	95.852	ug/l	86
9) 1,1,2-Trichlorotrifluo...	4.202	101	784949	95.666	ug/l	96
10) Methyl Iodide	4.411	142	1210076	99.948	ug/l	97
11) Tert butyl alcohol	5.388	59	1189022	502.601	ug/l	98
12) 1,1-Dichloroethene	4.173	96	805015	96.040	ug/l	93
13) Acrolein	4.036	56	1255019	481.856	ug/l	100
14) Allyl chloride	4.832	41	1880118	98.461	ug/l #	84
15) Acrylonitrile	5.551	53	3245966	491.326	ug/l	99
16) Acetone	4.285	43	3061853	454.819	ug/l	97
17) Carbon Disulfide	4.521	76	2470198	94.891	ug/l	99
18) Methyl Acetate	4.851	43	1480417	96.905	ug/l	93
19) Methyl tert-butyl Ether	5.616	73	3094606	96.697	ug/l	95
20) Methylene Chloride	5.090	84	950788	93.004	ug/l	89
21) trans-1,2-Dichloroethene	5.591	96	857004	95.894	ug/l	89
22) Diisopropyl ether	6.495	45	3507347	94.828	ug/l	92
23) Vinyl Acetate	6.431	43	15928569	486.887	ug/l	96
24) 1,1-Dichloroethane	6.383	63	1772865	95.197	ug/l	98
25) 2-Butanone	7.335	43	4724869	486.229	ug/l	93
26) 2,2-Dichloropropane	7.324	77	1401420	95.819	ug/l	98
27) cis-1,2-Dichloroethene	7.324	96	1009601	95.650	ug/l	88
28) Bromochloromethane	7.657	49	778999	95.860	ug/l #	79
29) Tetrahydrofuran	7.683	42	3073640	486.958	ug/l	89
30) Chloroform	7.818	83	1654382	93.380	ug/l	100
31) Cyclohexane	8.094	56	1758036	91.905	ug/l	87
32) 1,1,1-Trichloroethane	8.013	97	1442474	96.458	ug/l	98
36) 1,1-Dichloropropene	8.220	75	1296072	96.258	ug/l	97
37) Ethyl Acetate	7.413	43	1777883	95.805	ug/l	99
38) Carbon Tetrachloride	8.204	117	1204251	97.936	ug/l	98
39) Methylcyclohexane	9.459	83	1649325	99.643	ug/l	90
40) Benzene	8.459	78	3946754	93.300	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070893.D
 Acq On : 2 Feb 2022 14:37
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Feb 03 02:42:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.635	41	916792	98.326	ug/l #	88
42) 1,2-Dichloroethane	8.528	62	1355761	94.145	ug/l	100
43) Isopropyl Acetate	8.560	43	2849119	91.329	ug/l	95
44) Trichloroethene	9.215	130	938454	95.874	ug/l	93
45) 1,2-Dichloropropane	9.488	63	1082091	95.527	ug/l	99
46) Dibromomethane	9.577	93	648856	95.620	ug/l	96
47) Bromodichloromethane	9.762	83	1333262	97.543	ug/l	99
48) Methyl methacrylate	9.558	41	1390202	103.359	ug/l #	79
49) 1,4-Dioxane	9.569	88	454450	2016.921	ug/l #	84
51) 4-Methyl-2-Pentanone	10.331	43	9115898	496.447	ug/l	99
52) Toluene	10.505	92	2424895	92.517	ug/l	97
53) t-1,3-Dichloropropene	10.717	75	1560589	101.168	ug/l	99
54) cis-1,3-Dichloropropene	10.188	75	1678091	100.021	ug/l	99
55) 1,1,2-Trichloroethane	10.896	97	948891	95.326	ug/l	98
56) Ethyl methacrylate	10.760	69	1776565	102.469	ug/l	89
57) 1,3-Dichloropropane	11.044	76	1685082	96.699	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.041	63	2341351	479.178	ug/l	91
59) 2-Hexanone	11.084	43	6816809	508.470	ug/l	98
60) Dibromochloromethane	11.237	129	988423	100.115	ug/l	100
61) 1,2-Dibromoethane	11.344	107	1000383	97.718	ug/l	99
64) Tetrachloroethene	10.977	164	784399	91.682	ug/l	96
65) Chlorobenzene	11.768	112	2543501	92.970	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.840	131	914865	95.810	ug/l	99
67) Ethyl Benzene	11.843	91	4792154	96.088	ug/l	97
68) m/p-Xylenes	11.950	106	3567150	193.039	ug/l	95
69) o-Xylene	12.278	106	1779446	96.216	ug/l	93
70) Styrene	12.291	104	3005469	103.111	ug/l	97
71) Bromoform	12.457	173	793427	100.636	ug/l #	99
73) Isopropylbenzene	12.578	105	4610108	92.861	ug/l	98
74) N-amyl acetate	12.388	43	2187561	98.388	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.825	83	1515393	88.124	ug/l	99
76) 1,2,3-Trichloropropane	12.878	75	1952795	142.584	ug/l	72
77) Bromobenzene	12.860	156	1089334	91.530	ug/l	79
78) n-propylbenzene	12.919	91	5432299	95.949	ug/l	97
79) 2-Chlorotoluene	13.007	91	3286501	93.353	ug/l	93
80) 1,3,5-Trimethylbenzene	13.058	105	3821898	94.968	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.624	75	520076	102.932	ug/l	89
82) 4-Chlorotoluene	13.104	91	3224118	95.022	ug/l	92
83) tert-Butylbenzene	13.321	119	3339919	95.161	ug/l	95
84) 1,2,4-Trimethylbenzene	13.367	105	3748533	95.250	ug/l	97
85) sec-Butylbenzene	13.501	105	4775638	97.255	ug/l	97
86) p-Isopropyltoluene	13.613	119	3811683	98.435	ug/l	97
87) 1,3-Dichlorobenzene	13.613	146	1849169	93.407	ug/l	97
88) 1,4-Dichlorobenzene	13.694	146	1811488	91.476	ug/l	98
89) n-Butylbenzene	13.940	91	3353867	98.282	ug/l	97
90) Hexachloroethane	14.209	117	682554	99.460	ug/l	96
91) 1,2-Dichlorobenzene	13.989	146	1741013	91.669	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.603	75	309942	99.466	ug/l	85
93) 1,2,4-Trichlorobenzene	15.263	180	1004096	95.539	ug/l	99
94) Hexachlorobutadiene	15.367	225	553244	87.222	ug/l	99
95) Naphthalene	15.504	128	2900703	93.870	ug/l	100
96) 1,2,3-Trichlorobenzene	15.700	180	966454	94.840	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
Data File : VN070893.D
Acq On : 2 Feb 2022 14:37
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Quant Time: Feb 03 02:42:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 03 02:42:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.086 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion:168 Resp: 831747

Ion Ratio Lower Upper

168 100

99 57.1 41.3 61.9



#2

Dichlorodifluoromethane

Concen: 95.952 ug/l

RT: 2.070 min Scan# 30

Delta R.T. -0.003 min

Lab File: VN070893.D




Acq: 2 Feb 2022 14:37


Tgt Ion: 85 Resp: 885515

Ion Ratio Lower Upper

85 100

87 32.5 16.3 48.9





#3

Chloromethane

Concen: 98.632 ug/l

RT: 2.301 min Scan# 117

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 50 Resp: 1149685

Ion Ratio Lower Upper

50 100

52 32.4 25.4 38.0



#4

Vinyl Chloride

Concen: 94.390 ug/l

RT: 2.443 min Scan# 169

Delta R.T. -0.003 min

Lab File: VN070893.D




Acq: 2 Feb 2022 14:37


Tgt Ion: 62 Resp: 919140

Ion Ratio Lower Upper

62 100

64 31.6 25.8 38.6









#5
Bromomethane
Concen: 102.168 ug/l
RT: 2.818 min Scan# 309
Delta R.T. -0.014 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 94 Resp: 508855
Ion Ratio Lower Upper
94 100
96 92.9 73.8 110.8



#6
Chloroethane
Concen: 95.493 ug/l
RT: 2.995 min Scan# 375
Delta R.T. -0.014 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 64 Resp: 421722
Ion Ratio Lower Upper
64 100
66 31.8 26.3 39.5









#7
 Trichlorofluoromethane
 Concen: 94.010 ug/l
 RT: 3.363 min Scan# 51
 Delta R.T. -0.008 min
 Lab File: VN070893.D
 Acq: 2 Feb 2022 14:37




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:101 Resp: 1022303
 Ion Ratio Lower Upper
 101 100
 103 64.6 50.1 75.1



#8
 Diethyl Ether
 Concen: 95.852 ug/l
 RT: 3.819 min Scan# 682
 Delta R.T. -0.002 min
 Lab File: VN070893.D
 Acq: 2 Feb 2022 14:37

Tgt Ion: 74 Resp: 604523
 Ion Ratio Lower Upper
 74 100
 45 107.8 47.0 141.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 95.666 ug/l
RT: 4.202 min Scan# 81
Delta R.T. -0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 784949
Ion Ratio Lower Upper
101 100
85 44.5 37.2 55.8
151 78.0 65.9 98.9



#10
Methyl Iodide
Concen: 99.948 ug/l
RT: 4.411 min Scan# 903
Delta R.T. -0.006 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion:142 Resp: 1210076
Ion Ratio Lower Upper
142 100
127 42.0 35.4 53.0
141 14.1 11.6 17.4









#11
Tert butyl alcohol
Concen: 502.601 ug/l
RT: 5.388 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 1189022
Ion Ratio Lower Upper
59 100
57 10.1 8.7 13.1



#12
1,1-Dichloroethene
Concen: 96.040 ug/l
RT: 4.173 min Scan# 814
Delta R.T. -0.002 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 96 Resp: 805015
Ion Ratio Lower Upper
96 100
61 175.6 131.0 196.6
98 64.1 52.4 78.6









#13
Acrolein
Concen: 481.856 ug/l
RT: 4.036 min Scan# 70
Delta R.T. -0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 1255019
Ion Ratio Lower Upper
56 100
55 70.8 56.6 85.0



#14
Allyl chloride
Concen: 98.461 ug/l
RT: 4.832 min Scan# 1060
Delta R.T. -0.006 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 41 Resp: 1880118
Ion Ratio Lower Upper
41 100
39 60.6 61.2 91.8#
76 27.6 26.0 39.0









#15
Acrylonitrile
Concen: 491.326 ug/l
RT: 5.551 min Scan# 1328
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 53 Resp: 3245966
Ion Ratio Lower Upper
53 100
52 83.4 66.2 99.2
51 35.7 29.6 44.4



#16
Acetone
Concen: 454.819 ug/l
RT: 4.285 min Scan# 856
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 43 Resp: 3061853
Ion Ratio Lower Upper
43 100
58 30.3 23.1 34.7









#17
Carbon Disulfide
Concen: 94.891 ug/l
RT: 4.521 min Scan# 94
Delta R.T. -0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 76 Resp: 2470198
Ion Ratio Lower Upper
76 100
78 9.0 7.5 11.3



#18
Methyl Acetate
Concen: 96.905 ug/l
RT: 4.851 min Scan# 1067
Delta R.T. -0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 43 Resp: 1480417
Ion Ratio Lower Upper
43 100
74 20.9 19.4 29.2





#19

Methyl tert-butyl Ether

Concen: 96.697 ug/l

RT: 5.616 min Scan# 1156

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 73 Resp: 3094606

Ion Ratio Lower Upper

73 100

57 23.2 16.8 25.2



#20

Methylene Chloride

Concen: 93.004 ug/l

RT: 5.090 min Scan# 1156

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Tgt Ion: 84 Resp: 950788




Ion Ratio Lower Upper


84 100

49 136.4 93.5 140.3

51 41.3 29.4 44.2

86 65.0 53.0 79.4









#21
trans-1,2-Dichloroethene
Concen: 95.894 ug/l
RT: 5.591 min Scan# 11
Delta R.T. -0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 857004
Ion Ratio Lower Upper
96 100
61 151.1 105.8 158.6
98 63.6 52.0 78.0



#22
Diisopropyl ether
Concen: 94.828 ug/l
RT: 6.495 min Scan# 1680
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 45 Resp: 3507347
Ion Ratio Lower Upper
45 100
43 57.4 51.7 77.5
87 23.9 23.2 34.8
59 10.9 9.4 14.0





#23

Vinyl Acetate

Concen: 486.887 ug/l

RT: 6.431 min Scan# 1

Delta R.T. -0.003 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp:15928569

Ion Ratio Lower Upper

43 100

86 9.0 8.5 12.7



#24

1,1-Dichloroethane

Concen: 95.197 ug/l

RT: 6.383 min Scan# 1638

Delta R.T. -0.002 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37




Tgt Ion: 63 Resp: 1772865


Ion Ratio Lower Upper

63 100

98 5.8 3.3 9.9

100 3.8 2.1 6.5





#25

2-Butanone

Concen: 486.229 ug/l

RT: 7.335 min Scan# 1992

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 4724869

Ion Ratio Lower Upper

43 100

72 23.5 21.7 32.5



#26

2,2-Dichloropropane

Concen: 95.819 ug/l

RT: 7.324 min Scan# 1989

Delta R.T. 0.003 min

Lab File: VN070893.D




Acq: 2 Feb 2022 14:37


Tgt Ion: 77 Resp: 1401420

Ion Ratio Lower Upper

77 100

97 22.4 11.6 34.8





#27

cis-1,2-Dichloroethene

Concen: 95.650 ug/l

RT: 7.324 min Scan# 1989

Delta R.T. 0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :





MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 96 Resp: 1009601

Ion	Ratio	Lower	Upper
96	100		
61	157.6	0.0	275.0
98	64.4	0.0	129.6



#28

Bromochloromethane

Concen: 95.860 ug/l

RT: 7.657 min Scan# 2113




Delta R.T. -0.000 min


Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Tgt Ion: 49 Resp: 778999

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.0
130	64.9	67.8	101.8





#29

Tetrahydrofuran

Concen: 486.958 ug/l

RT: 7.683 min Scan# 2123

Delta R.T. -0.003 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :





MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 42 Resp: 3073640

Ion	Ratio	Lower	Upper
42	100		
72	37.8	36.3	54.5
71	36.2	34.7	52.1



#30

Chloroform

Concen: 93.380 ug/l

RT: 7.818 min Scan# 2173




Delta R.T. -0.000 min


Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Tgt Ion: 83 Resp: 1654382

Ion	Ratio	Lower	Upper
83	100		
85	64.8	51.9	77.9









#31
Cyclohexane
Concen: 91.905 ug/l
RT: 8.094 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 1758036
Ion Ratio Lower Upper
56 100
69 30.7 27.1 40.7
84 76.5 73.4 110.0



#32
1,1,1-Trichloroethane
Concen: 96.458 ug/l
RT: 8.013 min Scan# 2246
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 97 Resp: 1442474
Ion Ratio Lower Upper
97 100
99 64.3 50.1 75.1
61 45.6 35.2 52.8





#33

1,2-Dichloroethane-d4

Concen: 91.853 ug/l

RT: 8.434 min Scan# 2403

Delta R.T. -0.003 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 65 Resp: 1034134

Ion Ratio Lower Upper

65 100

67 53.6 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.968 min Scan# 2602

Delta R.T. 0.003 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37




Tgt Ion: 114 Resp: 1383536


Ion Ratio Lower Upper

114 100

63 22.2 0.0 37.8

88 16.1 0.0 27.4









#35
Dibromofluoromethane
Concen: 92.494 ug/l
RT: 8.019 min Scan# 211
Delta R.T. -0.002 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:113 Resp: 766605
Ion Ratio Lower Upper
113 100
111 103.4 82.2 123.2
192 20.4 16.9 25.3



#36
1,1-Dichloropropene
Concen: 96.258 ug/l
RT: 8.220 min Scan# 2323
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 75 Resp: 1296072
Ion Ratio Lower Upper
75 100
110 33.4 18.4 55.2
77 30.3 24.4 36.6





#37

Ethyl Acetate

Concen: 95.805 ug/l

RT: 7.413 min Scan# 20

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :





MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 1777883

Ion	Ratio	Lower	Upper
43	100		
61	13.6	11.5	17.3
70	9.7	7.9	11.9



#38

Carbon Tetrachloride

Concen: 97.936 ug/l

RT: 8.204 min Scan# 2317




Delta R.T. -0.002 min


Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Tgt Ion: 117 Resp: 1204251

Ion	Ratio	Lower	Upper
117	100		
119	96.6	76.1	114.1
121	31.1	24.2	36.2





#39

Methylcyclohexane

Concen: 99.643 ug/l

RT: 9.459 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :





MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 83 Resp: 1649325

Ion	Ratio	Lower	Upper
83	100		
55	83.1	58.1	87.1
98	44.3	38.4	57.6



#40

Benzene

Concen: 93.300 ug/l

RT: 8.459 min Scan# 2412




Delta R.T. -0.000 min


Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Tgt Ion: 78 Resp: 3946754

Ion	Ratio	Lower	Upper
78	100		
77	23.7	18.6	27.8









#41
Methacrylonitrile
Concen: 98.326 ug/l
RT: 7.635 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 41 Resp: 916792
Ion Ratio Lower Upper
41 100
39 56.4 48.1 72.1
67 69.9 68.3 102.5
52 28.7 29.6 44.4#



#42
1,2-Dichloroethane
Concen: 94.145 ug/l
RT: 8.528 min Scan# 2438
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 62 Resp: 1355761
Ion Ratio Lower Upper
62 100
98 8.7 0.0 17.6









#43
Isopropyl Acetate
Concen: 91.329 ug/l
RT: 8.560 min Scan# 2450
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 43 Resp: 2849119




Ion	Ratio	Lower	Upper
43	100		
61	23.6	20.4	30.6
87	11.5	11.2	16.8




#44
Trichloroethene
Concen: 95.874 ug/l
RT: 9.215 min Scan# 2694
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 130 Resp: 938454

Ion	Ratio	Lower	Upper
130	100		
95	102.0	0.0	191.2





#45

1,2-Dichloropropane

Concen: 95.527 ug/l

RT: 9.488 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion: 63 Resp: 1082091

Ion Ratio Lower Upper

63 100

65 30.5 24.7 37.1



#46

Dibromomethane

Concen: 95.620 ug/l

RT: 9.577 min Scan# 2829

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37




Tgt Ion: 93 Resp: 648856


Ion Ratio Lower Upper

93 100

95 83.0 65.9 98.9

174 98.6 84.6 127.0









#47
Bromodichloromethane
Concen: 97.543 ug/l
RT: 9.762 min Scan# 2898
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 1333262
Ion Ratio Lower Upper
83 100
85 64.2 50.9 76.3
127 8.2 7.1 10.7



#48
Methyl methacrylate
Concen: 103.359 ug/l
RT: 9.558 min Scan# 2822
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 41 Resp: 1390202
Ion Ratio Lower Upper
41 100
69 75.5 72.4 108.6
39 49.8 58.4 87.6#





#49

1,4-Dioxane

Concen: 2016.921 ug/l

RT: 9.569 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion: 88 Resp: 454450

Ion Ratio Lower Upper

88 100

43 33.7 23.4 35.2

58 78.2 50.6 75.8#



#50

Toluene-d8

Concen: 92.285 ug/l

RT: 10.440 min Scan# 3151

Delta R.T. -0.001 min

Lab File: VN070893.D




Acq: 2 Feb 2022 14:37


Tgt Ion: 98 Resp: 3140469

Ion Ratio Lower Upper

98 100

100 65.0 50.4 75.6









#51
4-Methyl-2-Pentanone
Concen: 496.447 ug/l
RT: 10.331 min Scan# 3109
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 9115898
Ion Ratio Lower Upper
43 100
58 37.6 30.4 45.6



#52
Toluene
Concen: 92.517 ug/l
RT: 10.505 min Scan# 3175
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 92 Resp: 2424895
Ion Ratio Lower Upper
92 100
91 172.9 142.0 213.0









#53
t-1,3-Dichloropropene
Concen: 101.168 ug/l
RT: 10.717 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 75 Resp: 1560589
Ion Ratio Lower Upper
75 100
77 31.4 25.6 38.4



#54
cis-1,3-Dichloropropene
Concen: 100.021 ug/l
RT: 10.188 min Scan# 3057
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 75 Resp: 1678091
Ion Ratio Lower Upper
75 100
77 31.3 26.4 39.6
39 53.4 42.5 63.7









#55
1,1,2-Trichloroethane
Concen: 95.326 ug/l
RT: 10.896 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 97 Resp: 948891
Ion Ratio Lower Upper
97 100
83 88.4 67.8 101.6
85 57.5 45.4 68.2
99 61.9 49.4 74.0



#56
Ethyl methacrylate
Concen: 102.469 ug/l
RT: 10.760 min Scan# 3270
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 69 Resp: 1776565
Ion Ratio Lower Upper
69 100
41 72.6 51.4 77.0
39 35.9 34.6 52.0





#57

1,3-Dichloropropane

Concen: 96.699 ug/l

RT: 11.044 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion: 76 Resp: 1685082

Ion Ratio Lower Upper

76 100

78 32.0 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 479.178 ug/l

RT: 10.041 min Scan# 3002

Delta R.T. -0.000 min

Lab File: VN070893.D




Acq: 2 Feb 2022 14:37


Tgt Ion: 63 Resp: 2341351

Ion Ratio Lower Upper

63 100

106 25.1 23.9 35.9









#59
2-Hexanone
Concen: 508.470 ug/l
RT: 11.084 min Scan# 3391
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 6816809
Ion Ratio Lower Upper
43 100
58 52.3 26.9 80.7



#60
Dibromochloromethane
Concen: 100.115 ug/l
RT: 11.237 min Scan# 3448
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 129 Resp: 988423
Ion Ratio Lower Upper
129 100
127 77.1 38.6 116.0









#61
1,2-Dibromoethane
Concen: 97.718 ug/l
RT: 11.344 min Scan# 3488
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:107 Resp: 1000383
Ion Ratio Lower Upper
107 100
109 94.0 75.7 113.5



#62
4-Bromofluorobenzene
Concen: 98.450 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 95 Resp: 1207180
Ion Ratio Lower Upper
95 100
174 77.2 0.0 187.2
176 74.7 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.744 min Scan# 3
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion:117 Resp: 1307168
Ion Ratio Lower Upper
117 100
82 58.5 42.4 63.6
119 32.4 25.4 38.2



#64
Tetrachloroethene
Concen: 91.682 ug/l
RT: 10.977 min Scan# 3351
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion:164 Resp: 784399
Ion Ratio Lower Upper
164 100
166 128.5 101.9 152.9
129 97.1 72.0 108.0
131 93.8 70.9 106.3









#65
Chlorobenzene
Concen: 92.970 ug/l
RT: 11.768 min Scan# 3646
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:112 Resp: 2543501
Ion Ratio Lower Upper
112 100
114 32.0 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 95.810 ug/l
RT: 11.840 min Scan# 3673
Delta R.T. -0.001 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion:131 Resp: 914865
Ion Ratio Lower Upper
131 100
133 95.5 47.9 143.7
119 65.0 32.1 96.5









#67
Ethyl Benzene
Concen: 96.088 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 4792154
Ion Ratio Lower Upper
91 100
106 30.3 25.7 38.5



#68
m/p-Xylenes
Concen: 193.039 ug/l
RT: 11.950 min Scan# 3714
Delta R.T. -0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion:106 Resp: 3567150
Ion Ratio Lower Upper
106 100
91 208.9 160.6 240.8









#69
o-Xylene
Concen: 96.216 ug/l
RT: 12.278 min Scan# 3836
Delta R.T. -0.002 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:106 Resp: 1779446
Ion Ratio Lower Upper
106 100
91 222.0 105.8 317.3



#70
Styrene
Concen: 103.111 ug/l
RT: 12.291 min Scan# 3841
Delta R.T. -0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion:104 Resp: 3005469
Ion Ratio Lower Upper
104 100
78 52.5 38.7 58.1
103 55.4 45.0 67.6





#71

Bromoform

Concen: 100.636 ug/l

RT: 12.457 min Scan# 3902

Delta R.T. 0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:173 Resp: 793427

Ion Ratio Lower Upper

173 100

175 49.0 24.2 72.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.672 min Scan# 4356

Delta R.T. 0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37




Tgt Ion:152 Resp: 569003


Ion Ratio Lower Upper

152 100

115 60.1 27.5 82.5

150 188.6 0.0 344.4









#73
Isopropylbenzene
Concen: 92.861 ug/l
RT: 12.578 min Scan# 3947
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:105 Resp: 4610108
Ion Ratio Lower Upper
105 100
120 25.7 13.4 40.1



#74
N-ethyl acetate
Concen: 98.388 ug/l
RT: 12.388 min Scan# 3877
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 43 Resp: 2187561
Ion Ratio Lower Upper
43 100
70 38.0 33.4 50.0
55 22.3 19.3 28.9
61 22.4 18.5 27.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 88.124 ug/l
 RT: 12.825 min Scan# 4040
 Delta R.T. -0.000 min
 Lab File: VN070893.D
 Acq: 2 Feb 2022 14:37




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion: 83 Resp: 1515393
 Ion Ratio Lower Upper
 83 100
 131 9.7 5.5 16.5
 85 64.1 32.4 97.2



#76
 1,2,3-Trichloropropane
 Concen: 142.584 ug/l
 RT: 12.878 min Scan# 4060
 Delta R.T. 0.000 min
 Lab File: VN070893.D
 Acq: 2 Feb 2022 14:37

Tgt Ion: 75 Resp: 1952795
 Ion Ratio Lower Upper
 75 100
 77 127.0 82.6 247.8









#77
Bromobenzene
Concen: 91.530 ug/l
RT: 12.860 min Scan# 4053
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:156 Resp: 1089334
Ion Ratio Lower Upper
156 100
77 227.7 90.6 271.8
158 97.0 48.9 146.7



#78
n-propylbenzene
Concen: 95.949 ug/l
RT: 12.919 min Scan# 4075
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 91 Resp: 5432299
Ion Ratio Lower Upper
91 100
120 22.2 11.9 35.9









#79
2-Chlorotoluene
Concen: 93.353 ug/l
RT: 13.007 min Scan# 4107
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 3286501
Ion Ratio Lower Upper
91 100
126 32.2 18.3 54.8



#80
1,3,5-Trimethylbenzene
Concen: 94.968 ug/l
RT: 13.058 min Scan# 4127
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 105 Resp: 3821898
Ion Ratio Lower Upper
105 100
120 47.2 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 102.932 ug/l
RT: 12.624 min Scan# 3965
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 520076
Ion Ratio Lower Upper
75 100
53 98.9 90.5 135.7
89 44.6 38.2 57.4



#82
4-Chlorotoluene
Concen: 95.022 ug/l
RT: 13.104 min Scan# 4144
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 91 Resp: 3224118
Ion Ratio Lower Upper
91 100
126 31.9 18.2 54.6





#83

tert-Butylbenzene

Concen: 95.161 ug/l

RT: 13.321 min Scan# 41

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:119 Resp: 3339919

Ion Ratio Lower Upper

119 100

91 68.1 31.5 94.5

134 25.6 12.4 37.4



#84

1,2,4-Trimethylbenzene

Concen: 95.250 ug/l

RT: 13.367 min Scan# 4242

Delta R.T. -0.000 min

Lab File: VN070893.D




Acq: 2 Feb 2022 14:37


Tgt Ion:105 Resp: 3748533

Ion Ratio Lower Upper

105 100

120 43.9 22.8 68.4









#85
 sec-Butylbenzene
 Concen: 97.255 ug/l
 RT: 13.501 min Scan# 4292
 Delta R.T. -0.000 min
 Lab File: VN070893.D
 Acq: 2 Feb 2022 14:37




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:105 Resp: 4775638
 Ion Ratio Lower Upper
 105 100
 134 19.5 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 98.435 ug/l
 RT: 13.613 min Scan# 4334
 Delta R.T. 0.000 min
 Lab File: VN070893.D
 Acq: 2 Feb 2022 14:37

Tgt Ion:119 Resp: 3811683
 Ion Ratio Lower Upper
 119 100
 134 25.7 13.6 40.7
 91 24.5 11.4 34.1









#87
1,3-Dichlorobenzene
Concen: 93.407 ug/l
RT: 13.613 min Scan# 4334
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:146 Resp: 1849169
Ion Ratio Lower Upper
146 100
111 41.6 18.6 55.6
148 63.8 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 91.476 ug/l
RT: 13.694 min Scan# 4364
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion:146 Resp: 1811488
Ion Ratio Lower Upper
146 100
111 40.0 18.3 54.8
148 63.5 31.8 95.4









#89
n-Butylbenzene
Concen: 98.282 ug/l
RT: 13.940 min Scan# 4456
Delta R.T. -0.001 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 3353867
Ion Ratio Lower Upper
91 100
92 52.4 25.4 76.0
134 24.7 13.6 40.8



#90
Hexachloroethane
Concen: 99.460 ug/l
RT: 14.209 min Scan# 4556
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 117 Resp: 682554
Ion Ratio Lower Upper
117 100
201 84.2 43.9 131.6









#91
1,2-Dichlorobenzene
Concen: 91.669 ug/l
RT: 13.989 min Scan# 4473
Delta R.T. 0.003 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:146 Resp: 1741013
Ion Ratio Lower Upper
146 100
111 42.7 19.1 57.3
148 63.8 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 99.466 ug/l
RT: 14.603 min Scan# 4703
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion: 75 Resp: 309942
Ion Ratio Lower Upper
75 100
155 80.6 46.9 140.7
157 104.0 60.6 181.7





#93

1,2,4-Trichlorobenzene

Concen: 95.539 ug/l

RT: 15.263 min Scan# 4949

Delta R.T. -0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:180 Resp: 1004096

Ion Ratio Lower Upper

180 100

182 95.5 48.3 144.9

145 30.2 14.5 43.5



#94

Hexachlorobutadiene

Concen: 87.222 ug/l

RT: 15.367 min Scan# 4988

Delta R.T. 0.000 min

Lab File: VN070893.D

Acq: 2 Feb 2022 14:37




Tgt Ion:225 Resp: 553244


Ion Ratio Lower Upper

225 100

223 62.8 31.9 95.7

227 63.2 32.1 96.3









#95
Naphthalene
Concen: 93.870 ug/l
RT: 15.504 min Scan# 5039
Delta R.T. 0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:128 Resp: 2900703
Ion Ratio Lower Upper
128 100
127 13.0 10.5 15.7
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 94.840 ug/l
RT: 15.700 min Scan# 5112
Delta R.T. -0.000 min
Lab File: VN070893.D
Acq: 2 Feb 2022 14:37

Tgt Ion:180 Resp: 966454
Ion Ratio Lower Upper
180 100
182 94.8 47.4 142.2
145 32.0 15.0 45.1

