

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070894.D
 Acq On : 2 Feb 2022 15:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Feb 03 02:43:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	812278	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1381604	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1327959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	560732	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	1667294	151.641	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 303.280%#	
35) Dibromofluoromethane	8.019	113	1233271	149.007	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 298.020%#	
50) Toluene-d8	10.440	98	5076648	149.389	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 298.780%#	
62) 4-Bromofluorobenzene	12.731	95	1973764	161.193	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 322.380%#	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.070	85	1358758	150.760	ug/l	99
3) Chloromethane	2.298	50	1804644	158.533	ug/l	98
4) Vinyl Chloride	2.445	62	1515046	159.315	ug/l	99
5) Bromomethane	2.818	94	742854	152.726	ug/l	100
6) Chloroethane	2.982	64	662502	153.610	ug/l	99
7) Trichlorofluoromethane	3.357	101	1640271	154.454	ug/l	98
8) Diethyl Ether	3.819	74	931127	151.177	ug/l	85
9) 1,1,2-Trichlorotrifluo...	4.194	101	1188999	148.383	ug/l	97
10) Methyl Iodide	4.409	142	1864223	157.669	ug/l	98
11) Tert butyl alcohol	5.396	59	1941382	840.294	ug/l	98
12) 1,1-Dichloroethene	4.167	96	1246473	152.271	ug/l	94
13) Acrolein	4.036	56	1892857	744.169	ug/l	99
14) Allyl chloride	4.830	41	2903036	155.675	ug/l #	85
15) Acrylonitrile	5.554	53	5049876	782.695	ug/l	99
16) Acetone	4.285	43	4680279	711.889	ug/l	98
17) Carbon Disulfide	4.519	76	3777888	148.603	ug/l	99
18) Methyl Acetate	4.851	43	2265053	152.089	ug/l	93
19) Methyl tert-butyl Ether	5.616	73	4758226	152.244	ug/l	94
20) Methylene Chloride	5.087	84	1460311	146.268	ug/l	90
21) trans-1,2-Dichloroethene	5.589	96	1293895	148.249	ug/l	88
22) Diisopropyl ether	6.495	45	5367291	148.593	ug/l	91
23) Vinyl Acetate	6.431	43	24329393	761.499	ug/l	96
24) 1,1-Dichloroethane	6.383	63	2742579	150.797	ug/l	98
25) 2-Butanone	7.335	43	7345140	773.995	ug/l	94
26) 2,2-Dichloropropane	7.321	77	2179128	152.563	ug/l	98
27) cis-1,2-Dichloroethene	7.321	96	1563905	151.717	ug/l	89
28) Bromochloromethane	7.654	49	1196080	150.712	ug/l #	79
29) Tetrahydrofuran	7.683	42	4746349	769.989	ug/l	89
30) Chloroform	7.815	83	2537992	146.687	ug/l	99
31) Cyclohexane	8.094	56	2682912	143.616	ug/l	86
32) 1,1,1-Trichloroethane	8.013	97	2228676	152.603	ug/l	98
36) 1,1-Dichloropropene	8.220	75	2009042	149.418	ug/l	97
37) Ethyl Acetate	7.412	43	2791933	150.659	ug/l	98
38) Carbon Tetrachloride	8.204	117	1869123	152.219	ug/l	99
39) Methylcyclohexane	9.459	83	2581646	156.187	ug/l	90
40) Benzene	8.458	78	6095494	144.297	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070894.D
 Acq On : 2 Feb 2022 15:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Feb 03 02:43:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.638	41	1427232	153.285	ug/l #	87
42) 1,2-Dichloroethane	8.528	62	2106527	146.483	ug/l	99
43) Isopropyl Acetate	8.560	43	4528583	145.367	ug/l	96
44) Trichloroethene	9.215	130	1445041	147.834	ug/l	91
45) 1,2-Dichloropropane	9.488	63	1676927	148.246	ug/l	100
46) Dibromomethane	9.577	93	1001207	147.751	ug/l	96
47) Bromodichloromethane	9.762	83	2094958	153.484	ug/l	99
48) Methyl methacrylate	9.561	41	2077301	154.659	ug/l #	83
49) 1,4-Dioxane	9.571	88	729964	3244.226	ug/l #	85
51) 4-Methyl-2-Pentanone	10.333	43	14442292	787.619	ug/l	100
52) Toluene	10.505	92	3758936	143.615	ug/l	96
53) t-1,3-Dichloropropene	10.717	75	2497851	162.154	ug/l	99
54) cis-1,3-Dichloropropene	10.188	75	2643566	157.788	ug/l	99
55) 1,1,2-Trichloroethane	10.899	97	1487129	149.606	ug/l	97
56) Ethyl methacrylate	10.760	69	2867496	165.623	ug/l	89
57) 1,3-Dichloropropane	11.044	76	2640171	151.719	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.041	63	3830408	772.481	ug/l	91
59) 2-Hexanone	11.087	43	10563924	789.071	ug/l	99
60) Dibromochloromethane	11.237	129	1572617	159.510	ug/l	100
61) 1,2-Dibromoethane	11.344	107	1593257	155.848	ug/l	99
64) Tetrachloroethene	10.977	164	1207743	138.953	ug/l	95
65) Chlorobenzene	11.768	112	4030508	145.016	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.840	131	1447150	149.182	ug/l	99
67) Ethyl Benzene	11.843	91	7513406	148.293	ug/l	97
68) m/p-Xylenes	11.953	106	5613382	299.017	ug/l	95
69) o-Xylene	12.278	106	2824108	150.311	ug/l	94
70) Styrene	12.294	104	4770298	161.096	ug/l	96
71) Bromoform	12.457	173	1253820	156.541	ug/l #	100
73) Isopropylbenzene	12.578	105	7294766	149.105	ug/l	98
74) N-amyl acetate	12.388	43	3454803	157.676	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.827	83	2272758	134.117	ug/l	99
76) 1,2,3-Trichloropropane	12.881	75	3003898	222.565	ug/l #	1
77) Bromobenzene	12.860	156	1691675	144.238	ug/l	79
78) n-propylbenzene	12.919	91	8594269	154.037	ug/l	97
79) 2-Chlorotoluene	13.007	91	5153242	148.537	ug/l	93
80) 1,3,5-Trimethylbenzene	13.058	105	5982741	150.854	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.626	75	819917	164.670	ug/l	89
82) 4-Chlorotoluene	13.104	91	5006861	149.741	ug/l	92
83) tert-Butylbenzene	13.324	119	5193831	150.165	ug/l	94
84) 1,2,4-Trimethylbenzene	13.366	105	5771134	148.807	ug/l	97
85) sec-Butylbenzene	13.501	105	7410222	153.133	ug/l	97
86) p-Isopropyltoluene	13.613	119	5857882	153.509	ug/l	97
87) 1,3-Dichlorobenzene	13.613	146	2836472	145.391	ug/l	97
88) 1,4-Dichlorobenzene	13.694	146	2793430	143.142	ug/l	98
89) n-Butylbenzene	13.940	91	5228661	155.481	ug/l	97
90) Hexachloroethane	14.209	117	1058295	156.487	ug/l	95
91) 1,2-Dichlorobenzene	13.986	146	2687972	143.617	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.603	75	502451	163.625	ug/l	85
93) 1,2,4-Trichlorobenzene	15.263	180	1622358	156.643	ug/l	98
94) Hexachlorobutadiene	15.367	225	840466	134.458	ug/l	99
95) Naphthalene	15.504	128	4801152	155.254	ug/l	100
96) 1,2,3-Trichlorobenzene	15.700	180	1555708	153.839	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
Data File : VN070894.D
Acq On : 2 Feb 2022 15:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Quant Time: Feb 03 02:43:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.083 min Scan# 21
Delta R.T. -0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:168 Resp: 812278
Ion Ratio Lower Upper
168 100
99 57.8 41.3 61.9



#2
Dichlorodifluoromethane
Concen: 150.760 ug/l
RT: 2.070 min Scan# 30
Delta R.T. -0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 85 Resp: 1358758
Ion Ratio Lower Upper
85 100
87 32.2 16.3 48.9









#3
Chloromethane
Concen: 158.533 ug/l
RT: 2.298 min Scan# 117
Delta R.T. -0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 50 Resp: 1804644
Ion Ratio Lower Upper
50 100
52 32.7 25.4 38.0



#4
Vinyl Chloride
Concen: 159.315 ug/l
RT: 2.445 min Scan# 170
Delta R.T. -0.001 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 62 Resp: 1515046
Ion Ratio Lower Upper
62 100
64 31.8 25.8 38.6









#5
Bromomethane
Concen: 152.726 ug/l
RT: 2.818 min Scan# 309
Delta R.T. -0.014 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 94 Resp: 742854
Ion Ratio Lower Upper
94 100
96 91.9 73.8 110.8



#6
Chloroethane
Concen: 153.610 ug/l
RT: 2.982 min Scan# 370
Delta R.T. -0.027 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 64 Resp: 662502
Ion Ratio Lower Upper
64 100
66 32.4 26.3 39.5









#7
Trichlorofluoromethane
Concen: 154.454 ug/l
RT: 3.357 min Scan# 51
Delta R.T. -0.014 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 1640271
Ion Ratio Lower Upper
101 100
103 64.5 50.1 75.1



#8
Diethyl Ether
Concen: 151.177 ug/l
RT: 3.819 min Scan# 682
Delta R.T. -0.002 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 74 Resp: 931127
Ion Ratio Lower Upper
74 100
45 108.9 47.0 141.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 148.383 ug/l
RT: 4.194 min Scan# 81
Delta R.T. -0.011 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 1188999
Ion Ratio Lower Upper
101 100
85 45.2 37.2 55.8
151 78.5 65.9 98.9



#10
Methyl Iodide
Concen: 157.669 ug/l
RT: 4.409 min Scan# 902
Delta R.T. -0.008 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:142 Resp: 1864223
Ion Ratio Lower Upper
142 100
127 42.2 35.4 53.0
141 14.4 11.6 17.4









#11
Tert butyl alcohol
Concen: 840.294 ug/l
RT: 5.396 min Scan# 11
Delta R.T. 0.014 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 59 Resp: 1941382
Ion Ratio Lower Upper
59 100
57 10.1 8.7 13.1



#12
1,1-Dichloroethene
Concen: 152.271 ug/l
RT: 4.167 min Scan# 812
Delta R.T. -0.008 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 96 Resp: 1246473
Ion Ratio Lower Upper
96 100
61 173.3 131.0 196.6
98 63.4 52.4 78.6









#13
Acrolein
Concen: 744.169 ug/l
RT: 4.036 min Scan# 70
Delta R.T. -0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 1892857
Ion Ratio Lower Upper
56 100
55 71.6 56.6 85.0



#14
Allyl chloride
Concen: 155.675 ug/l
RT: 4.830 min Scan# 1059
Delta R.T. -0.008 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 41 Resp: 2903036
Ion Ratio Lower Upper
41 100
39 60.8 61.2 91.8#
76 27.7 26.0 39.0





#15

Acrylonitrile

Concen: 782.695 ug/l

RT: 5.554 min Scan# 1

Delta R.T. 0.003 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :





MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 53 Resp: 5049876

Ion	Ratio	Lower	Upper
53	100		
52	83.5	66.2	99.2
51	35.6	29.6	44.4



#16

Acetone

Concen: 711.889 ug/l

RT: 4.285 min Scan# 856




Delta R.T. 0.000 min


Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Tgt Ion: 43 Resp: 4680279

Ion	Ratio	Lower	Upper
43	100		
58	30.1	23.1	34.7









#17
Carbon Disulfide
Concen: 148.603 ug/l
RT: 4.519 min Scan# 947
Delta R.T. -0.005 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 76 Resp: 3777888
Ion Ratio Lower Upper
76 100
78 9.2 7.5 11.3



#18
Methyl Acetate
Concen: 152.089 ug/l
RT: 4.851 min Scan# 1067
Delta R.T. -0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 43 Resp: 2265053
Ion Ratio Lower Upper
43 100
74 20.8 19.4 29.2





#19

Methyl tert-butyl Ether

Concen: 152.244 ug/l

RT: 5.616 min Scan# 1155

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 73 Resp: 4758226

Ion Ratio Lower Upper

73 100

57 23.7 16.8 25.2



#20

Methylene Chloride

Concen: 146.268 ug/l

RT: 5.087 min Scan# 1155

Delta R.T. -0.003 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Tgt Ion: 84 Resp: 1460311




Ion Ratio Lower Upper

84 100

49 134.4 93.5 140.3

51 41.1 29.4 44.2

86 64.5 53.0 79.4









#21
trans-1,2-Dichloroethene
Concen: 148.249 ug/l
RT: 5.589 min Scan# 11
Delta R.T. -0.005 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 1293895
Ion Ratio Lower Upper
96 100
61 152.1 105.8 158.6
98 63.4 52.0 78.0



#22
Diisopropyl ether
Concen: 148.593 ug/l
RT: 6.495 min Scan# 1680
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 45 Resp: 5367291
Ion Ratio Lower Upper
45 100
43 56.0 51.7 77.5
87 24.0 23.2 34.8
59 11.0 9.4 14.0









#23
 Vinyl Acetate
 Concen: 761.499 ug/l
 RT: 6.431 min Scan# 1656
 Delta R.T. -0.003 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion: 43 Resp:24329393
 Ion Ratio Lower Upper
 43 100
 86 9.2 8.5 12.7



#24
 1,1-Dichloroethane
 Concen: 150.797 ug/l
 RT: 6.383 min Scan# 1638
 Delta R.T. -0.002 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02

Tgt Ion: 63 Resp: 2742579
 Ion Ratio Lower Upper
 63 100
 98 6.0 3.3 9.9
 100 3.8 2.1 6.5





#25

2-Butanone

Concen: 773.995 ug/l

RT: 7.335 min Scan# 1992

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 43 Resp: 7345140

Ion Ratio Lower Upper

43 100

72 23.7 21.7 32.5



#26

2,2-Dichloropropane

Concen: 152.563 ug/l

RT: 7.321 min Scan# 1988

Delta R.T. 0.000 min

Lab File: VN070894.D




Acq: 2 Feb 2022 15:02


Tgt Ion: 77 Resp: 2179128

Ion Ratio Lower Upper

77 100

97 22.3 11.6 34.8





#27

cis-1,2-Dichloroethene

Concen: 151.717 ug/l

RT: 7.321 min Scan# 1989

Delta R.T. -0.003 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion: 96 Resp: 1563905

Ion Ratio Lower Upper

96 100

61 157.2 0.0 275.0

98 64.0 0.0 129.6



#28

Bromochloromethane

Concen: 150.712 ug/l

RT: 7.654 min Scan# 2112

Delta R.T. -0.003 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02




Tgt Ion: 49 Resp: 1196080

Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 65.0 67.8 101.8





#29

Tetrahydrofuran

Concen: 769.989 ug/l

RT: 7.683 min Scan# 2123

Delta R.T. -0.003 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion: 42 Resp: 4746349

Ion Ratio Lower Upper

42 100

72 38.1 36.3 54.5

71 36.2 34.7 52.1



#30

Chloroform

Concen: 146.687 ug/l

RT: 7.815 min Scan# 2172

Delta R.T. -0.003 min

Lab File: VN070894.D




Acq: 2 Feb 2022 15:02


Tgt Ion: 83 Resp: 2537992

Ion Ratio Lower Upper

83 100

85 64.4 51.9 77.9









#31
Cyclohexane
Concen: 143.616 ug/l
RT: 8.094 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 2682912
Ion Ratio Lower Upper
56 100
69 29.7 27.1 40.7
84 76.3 73.4 110.0



#32
1,1,1-Trichloroethane
Concen: 152.603 ug/l
RT: 8.013 min Scan# 2246
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 97 Resp: 2228676
Ion Ratio Lower Upper
97 100
99 63.8 50.1 75.1
61 45.6 35.2 52.8





#33

1,2-Dichloroethane-d4

Concen: 151.641 ug/l

RT: 8.434 min Scan# 2403

Delta R.T. -0.003 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 65 Resp: 1667294

Ion Ratio Lower Upper

65 100

67 53.6 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.965 min Scan# 2601

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02




Tgt Ion: 114 Resp: 1381604


Ion Ratio Lower Upper

114 100

63 23.5 0.0 37.8

88 16.7 0.0 27.4









#35
Dibromofluoromethane
Concen: 149.007 ug/l
RT: 8.019 min Scan# 21
Delta R.T. -0.002 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:113 Resp: 1233271
Ion Ratio Lower Upper
113 100
111 103.6 82.2 123.2
192 20.2 16.9 25.3



#36
1,1-Dichloropropene
Concen: 149.418 ug/l
RT: 8.220 min Scan# 2323
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 75 Resp: 2009042
Ion Ratio Lower Upper
75 100
110 33.3 18.4 55.2
77 30.6 24.4 36.6





#37

Ethyl Acetate

Concen: 150.659 ug/l

RT: 7.412 min Scan# 20

Delta R.T. -0.001 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :





MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 2791933

Ion	Ratio	Lower	Upper
43	100		
61	13.6	11.5	17.3
70	9.5	7.9	11.9



#38

Carbon Tetrachloride

Concen: 152.219 ug/l

RT: 8.204 min Scan# 2317




Delta R.T. -0.002 min


Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Tgt Ion: 117 Resp: 1869123

Ion	Ratio	Lower	Upper
117	100		
119	95.6	76.1	114.1
121	30.6	24.2	36.2









#39
Methylcyclohexane
Concen: 156.187 ug/l
RT: 9.459 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 2581646
Ion Ratio Lower Upper
83 100
55 82.8 58.1 87.1
98 43.9 38.4 57.6



#40
Benzene
Concen: 144.297 ug/l
RT: 8.458 min Scan# 2412
Delta R.T. -0.001 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02


Tgt Ion: 78 Resp: 6095494
Ion Ratio Lower Upper
78 100
77 23.7 18.6 27.8








#41
Methacrylonitrile
Concen: 153.285 ug/l
RT: 7.638 min Scan# 2105
Delta R.T. 0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150



Tgt Ion: 41 Resp: 1427232




Ion	Ratio	Lower	Upper
41	100		
39	55.9	48.1	72.1
67	69.5	68.3	102.5
52	29.2	29.6	44.4




#42
1,2-Dichloroethane
Concen: 146.483 ug/l
RT: 8.528 min Scan# 2438
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 62 Resp: 2106527

Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	17.6









#43
Isopropyl Acetate
Concen: 145.367 ug/l
RT: 8.560 min Scan# 2450
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion: 43 Resp: 4528583




Ion	Ratio	Lower	Upper
43	100		
61	23.7	20.4	30.6
87	11.5	11.2	16.8



#44
Trichloroethene
Concen: 147.834 ug/l
RT: 9.215 min Scan# 2694
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 130 Resp: 1445041

Ion	Ratio	Lower	Upper
130	100		
95	104.2	0.0	191.2





#45

1,2-Dichloropropane

Concen: 148.246 ug/l

RT: 9.488 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 63 Resp: 1676927

Ion Ratio Lower Upper

63 100

65 30.9 24.7 37.1



#46

Dibromomethane

Concen: 147.751 ug/l

RT: 9.577 min Scan# 2829

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02




Tgt Ion: 93 Resp: 1001207


Ion Ratio Lower Upper

93 100

95 83.1 65.9 98.9

174 99.0 84.6 127.0









#47
Bromodichloromethane
Concen: 153.484 ug/l
RT: 9.762 min Scan# 2898
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 2094958
Ion Ratio Lower Upper
83 100
85 64.1 50.9 76.3
127 8.1 7.1 10.7



#48
Methyl methacrylate
Concen: 154.659 ug/l
RT: 9.561 min Scan# 2823
Delta R.T. 0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 41 Resp: 2077301
Ion Ratio Lower Upper
41 100
69 79.7 72.4 108.6
39 53.1 58.4 87.6#









#49
1,4-Dioxane
Concen: 3244.226 ug/l
RT: 9.571 min Scan# 2827
Delta R.T. 0.002 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 88 Resp: 729964
Ion Ratio Lower Upper
88 100
43 33.4 23.4 35.2
58 77.8 50.6 75.8#



#50
Toluene-d8
Concen: 149.389 ug/l
RT: 10.440 min Scan# 3151
Delta R.T. -0.001 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 98 Resp: 5076648
Ion Ratio Lower Upper
98 100
100 65.0 50.4 75.6









#51
4-Methyl-2-Pentanone
Concen: 787.619 ug/l
RT: 10.333 min Scan# 3111
Delta R.T. 0.002 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp:14442292
Ion Ratio Lower Upper
43 100
58 38.2 30.4 45.6



#52
Toluene
Concen: 143.615 ug/l
RT: 10.505 min Scan# 3175
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 92 Resp: 3758936
Ion Ratio Lower Upper
92 100
91 172.5 142.0 213.0









#53
t-1,3-Dichloropropene
Concen: 162.154 ug/l
RT: 10.717 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 2497851
Ion Ratio Lower Upper
75 100
77 31.5 25.6 38.4



#54
cis-1,3-Dichloropropene
Concen: 157.788 ug/l
RT: 10.188 min Scan# 3057
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 75 Resp: 2643566
Ion Ratio Lower Upper
75 100
77 31.5 26.4 39.6
39 52.7 42.5 63.7









#55
1,1,2-Trichloroethane
Concen: 149.606 ug/l
RT: 10.899 min Scan# 31
Delta R.T. 0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion: 97 Resp: 1487129




Ion	Ratio	Lower	Upper
97	100		
83	89.7	67.8	101.6
85	57.1	45.4	68.2
99	62.6	49.4	74.0



#56
Ethyl methacrylate
Concen: 165.623 ug/l
RT: 10.760 min Scan# 3270
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 69 Resp: 2867496

Ion	Ratio	Lower	Upper
69	100		
41	73.1	51.4	77.0
39	36.3	34.6	52.0





#57

1,3-Dichloropropane

Concen: 151.719 ug/l

RT: 11.044 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 2640171

Ion Ratio Lower Upper

76 100

78 32.2 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 772.481 ug/l

RT: 10.041 min Scan# 3002

Delta R.T. -0.000 min

Lab File: VN070894.D




Acq: 2 Feb 2022 15:02


Tgt Ion: 63 Resp: 3830408

Ion Ratio Lower Upper

63 100

106 25.1 23.9 35.9









#59
2-Hexanone
Concen: 789.071 ug/l
RT: 11.087 min Scan# 31
Delta R.T. 0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp:10563924
Ion Ratio Lower Upper
43 100
58 52.9 26.9 80.7



#60
Dibromochloromethane
Concen: 159.510 ug/l
RT: 11.237 min Scan# 3448
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:129 Resp: 1572617
Ion Ratio Lower Upper
129 100
127 77.6 38.6 116.0









#61
1,2-Dibromoethane
Concen: 155.848 ug/l
RT: 11.344 min Scan# 3488
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:107 Resp: 1593257
Ion Ratio Lower Upper
107 100
109 93.7 75.7 113.5



#62
4-Bromofluorobenzene
Concen: 161.193 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 95 Resp: 1973764
Ion Ratio Lower Upper
95 100
174 76.3 0.0 187.2
176 74.2 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.744 min Scan# 3637
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 1327959
Ion Ratio Lower Upper
117 100
82 59.0 42.4 63.6
119 31.9 25.4 38.2



#64
Tetrachloroethene
Concen: 138.953 ug/l
RT: 10.977 min Scan# 3351
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:164 Resp: 1207743
Ion Ratio Lower Upper
164 100
166 128.4 101.9 152.9
129 98.7 72.0 108.0
131 93.5 70.9 106.3









#65
Chlorobenzene
Concen: 145.016 ug/l
RT: 11.768 min Scan# 3646
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:112 Resp: 4030508
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 149.182 ug/l
RT: 11.840 min Scan# 3673
Delta R.T. -0.001 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:131 Resp: 1447150
Ion Ratio Lower Upper
131 100
133 94.9 47.9 143.7
119 65.7 32.1 96.5









#67
Ethyl Benzene
Concen: 148.293 ug/l
RT: 11.843 min Scan# 301
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 7513406
Ion Ratio Lower Upper
91 100
106 30.6 25.7 38.5



#68
m/p-Xylenes
Concen: 299.017 ug/l
RT: 11.953 min Scan# 3715
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:106 Resp: 5613382
Ion Ratio Lower Upper
106 100
91 208.3 160.6 240.8





#69

o-Xylene

Concen: 150.311 ug/l

RT: 12.278 min Scan# 3836

Delta R.T. -0.002 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC150

Tgt Ion:106 Resp: 2824108

Ion Ratio Lower Upper

106 100

91 221.3 105.8 317.3



#70

Styrene

Concen: 161.096 ug/l

RT: 12.294 min Scan# 3842

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02




Tgt Ion:104 Resp: 4770298


Ion Ratio Lower Upper

104 100

78 52.8 38.7 58.1

103 55.5 45.0 67.6









#71
Bromoform
Concen: 156.541 ug/l
RT: 12.457 min Scan# 3902
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:173 Resp: 1253820
Ion Ratio Lower Upper
173 100
175 48.3 24.2 72.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.672 min Scan# 4356
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:152 Resp: 560732
Ion Ratio Lower Upper
152 100
115 61.3 27.5 82.5
150 207.6 0.0 344.4









#73
Isopropylbenzene
Concen: 149.105 ug/l
RT: 12.578 min Scan# 3947
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:105 Resp: 7294766
Ion Ratio Lower Upper
105 100
120 25.9 13.4 40.1



#74
N-ethyl acetate
Concen: 157.676 ug/l
RT: 12.388 min Scan# 3877
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 43 Resp: 3454803
Ion Ratio Lower Upper
43 100
70 38.0 33.4 50.0
55 22.2 19.3 28.9
61 22.6 18.5 27.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 134.117 ug/l
 RT: 12.827 min Scan# 4040
 Delta R.T. 0.002 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion: 83 Resp: 2272758
 Ion Ratio Lower Upper
 83 100
 131 9.7 5.5 16.5
 85 64.2 32.4 97.2



#76
 1,2,3-Trichloropropane
 Concen: 222.565 ug/l
 RT: 12.881 min Scan# 4061
 Delta R.T. 0.003 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02

Tgt Ion: 75 Resp: 3003898
 Ion Ratio Lower Upper
 75 100
 77 0.0 82.6 247.8#









#77
Bromobenzene
Concen: 144.238 ug/l
RT: 12.860 min Scan# 4053
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:156 Resp: 1691675
Ion Ratio Lower Upper
156 100
77 227.9 90.6 271.8
158 97.3 48.9 146.7



#78
n-propylbenzene
Concen: 154.037 ug/l
RT: 12.919 min Scan# 4075
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 91 Resp: 8594269
Ion Ratio Lower Upper
91 100
120 22.2 11.9 35.9









#79
2-Chlorotoluene
Concen: 148.537 ug/l
RT: 13.007 min Scan# 4108
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150

Tgt Ion: 91 Resp: 5153242
Ion Ratio Lower Upper
91 100
126 32.3 18.3 54.8



#80
1,3,5-Trimethylbenzene
Concen: 150.854 ug/l
RT: 13.058 min Scan# 4127
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 105 Resp: 5982741
Ion Ratio Lower Upper
105 100
120 47.5 24.6 73.6









#81
trans-1,4-Dichloro-2-butene
Concen: 164.670 ug/l
RT: 12.626 min Scan# 3965
Delta R.T. 0.002 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 819917
Ion Ratio Lower Upper
75 100
53 98.0 90.5 135.7
89 44.8 38.2 57.4



#82
4-Chlorotoluene
Concen: 149.741 ug/l
RT: 13.104 min Scan# 4144
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 91 Resp: 5006861
Ion Ratio Lower Upper
91 100
126 31.9 18.2 54.6









#83
tert-Butylbenzene
Concen: 150.165 ug/l
RT: 13.324 min Scan# 41
Delta R.T. 0.003 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:119 Resp: 5193831
Ion Ratio Lower Upper
119 100
91 68.6 31.5 94.5
134 25.5 12.4 37.4



#84
1,2,4-Trimethylbenzene
Concen: 148.807 ug/l
RT: 13.366 min Scan# 4242
Delta R.T. -0.001 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:105 Resp: 5771134
Ion Ratio Lower Upper
105 100
120 43.9 22.8 68.4









#85
 sec-Butylbenzene
 Concen: 153.133 ug/l
 RT: 13.501 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 7410222
 Ion Ratio Lower Upper
 105 100
 134 19.5 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 153.509 ug/l
 RT: 13.613 min Scan# 4334
 Delta R.T. 0.000 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02

Tgt Ion:119 Resp: 5857882
 Ion Ratio Lower Upper
 119 100
 134 25.8 13.6 40.7
 91 24.9 11.4 34.1









#87
1,3-Dichlorobenzene
Concen: 145.391 ug/l
RT: 13.613 min Scan# 4334
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion:146 Resp: 2836472
Ion Ratio Lower Upper
146 100
111 41.6 18.6 55.6
148 63.8 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 143.142 ug/l
RT: 13.694 min Scan# 4364
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:146 Resp: 2793430
Ion Ratio Lower Upper
146 100
111 39.9 18.3 54.8
148 63.6 31.8 95.4









#89
n-Butylbenzene
Concen: 155.481 ug/l
RT: 13.940 min Scan# 4456
Delta R.T. -0.001 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 5228661
Ion Ratio Lower Upper
91 100
92 52.6 25.4 76.0
134 24.8 13.6 40.8



#90
Hexachloroethane
Concen: 156.487 ug/l
RT: 14.209 min Scan# 4556
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion: 117 Resp: 1058295
Ion Ratio Lower Upper
117 100
201 83.3 43.9 131.6









#91
 1,2-Dichlorobenzene
 Concen: 143.617 ug/l
 RT: 13.986 min Scan# 4473
 Delta R.T. -0.000 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC150

Tgt Ion:146 Resp: 2687972




Ion	Ratio	Lower	Upper
146	100		
111	42.7	19.1	57.3
148	64.2	32.1	96.3




#92
 1,2-Dibromo-3-Chloropropane
 Concen: 163.625 ug/l
 RT: 14.603 min Scan# 4703
 Delta R.T. -0.000 min
 Lab File: VN070894.D
 Acq: 2 Feb 2022 15:02

Tgt Ion: 75 Resp: 502451

Ion	Ratio	Lower	Upper
75	100		
155	80.6	46.9	140.7
157	103.5	60.6	181.7









#93
1,2,4-Trichlorobenzene
Concen: 156.643 ug/l
RT: 15.263 min Scan# 4949
Delta R.T. -0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150


Tgt Ion:180 Resp: 1622358
Ion Ratio Lower Upper
180 100
182 95.1 48.3 144.9
145 30.7 14.5 43.5



#94
Hexachlorobutadiene
Concen: 134.458 ug/l
RT: 15.367 min Scan# 4988
Delta R.T. 0.000 min
Lab File: VN070894.D
Acq: 2 Feb 2022 15:02

Tgt Ion:225 Resp: 840466
Ion Ratio Lower Upper
225 100
223 62.6 31.9 95.7
227 63.4 32.1 96.3





#95

Naphthalene

Concen: 155.254 ug/l

RT: 15.504 min Scan# 5039

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150





Tgt Ion:128 Resp: 4801152

Ion Ratio Lower Upper

128 100

127 13.1 10.5 15.7

129 10.9 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 153.839 ug/l

RT: 15.700 min Scan# 5112

Delta R.T. -0.000 min

Lab File: VN070894.D

Acq: 2 Feb 2022 15:02

Tgt Ion:180 Resp: 1555708

Ion Ratio Lower Upper

180 100

182 95.4 47.4 142.2

145 32.3 15.0 45.1

