

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070897.D
 Acq On : 2 Feb 2022 17:12
 Operator : JC/MD
 Sample : VN0202MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0202MBL01

Quant Time: Feb 03 05:57:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1050690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1658416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1457216	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	559630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	646516	45.458	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.920%
35) Dibromofluoromethane	8.021	113	477880	48.101	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.200%
50) Toluene-d8	10.443	98	1989622	48.776	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.560%
62) 4-Bromofluorobenzene	12.731	95	669098	45.523	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.040%
Target Compounds						
93) 1,2,4-Trichlorobenzene	15.263	180	16748	1.620	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
Data File : VN070897.D
Acq On : 2 Feb 2022 17:12
Operator : JC/MD
Sample : VN0202MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0202MBL01

Quant Time: Feb 03 05:57:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

