

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020222\  
 Data File : VN070898.D  
 Acq On : 2 Feb 2022 17:43  
 Operator : JC/MD  
 Sample : N1303-03ME  
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 CC-6ME

Quant Time: Feb 02 18:02:01 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N020222W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Feb 02 15:30:39 2022  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|-----------------------------|----------------|------|------------|---------|--------|----------|
| -----                       |                |      |            |         |        |          |
| Internal Standards          |                |      |            |         |        |          |
| 1) Pentafluorobenzene       | 8.078          | 168  | 1167693    | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.963          | 114  | 1852125    | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.747         | 117  | 1450984    | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.675         | 152  | 635927     | 50.000  | ug/l   | 0.00     |
|                             |                |      |            |         |        |          |
| System Monitoring Compounds |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.432          | 65   | 730344     | 46.207  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 61 - 141 |      | Recovery = | 92.420% |        |          |
| 35) Dibromofluoromethane    | 8.019          | 113  | 527111     | 47.508  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 69 - 133 |      | Recovery = | 95.020% |        |          |
| 50) Toluene-d8              | 10.440         | 98   | 2156930    | 47.347  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 65 - 126 |      | Recovery = | 94.700% |        |          |
| 62) 4-Bromofluorobenzene    | 12.731         | 95   | 701863     | 42.758  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 58 - 135 |      | Recovery = | 85.520% |        |          |
|                             |                |      |            |         |        |          |
| Target Compounds            |                |      |            |         |        |          |
|                             |                |      |            |         | Qvalue |          |
| 5) Bromomethane             | 2.810          | 94   | 2955       | 0.423   | ug/l # | 80       |
| 18) Methyl Acetate          | 4.862          | 43   | 30181      | 0.939   | ug/l # | 86       |
| 29) Tetrahydrofuran         | 7.699          | 42   | 10544      | 1.190   | ug/l   | 91       |
| 31) Cyclohexane             | 8.078          | 56   | 24832      | 0.925   | ug/l # | 1        |
| 51) 4-Methyl-2-Pentanone    | 10.339         | 43   | 7330       | 0.298   | ug/l   | 96       |
| 59) 2-Hexanone              | 11.095         | 43   | 16544      | 0.922   | ug/l   | 91       |
| 74) N-amyl acetate          | 12.388         | 43   | 8999       | 0.362   | ug/l   | 93       |
| 84) 1,2,4-Trimethylbenzene  | 13.366         | 105  | 9175       | 0.209   | ug/l   | 98       |
| 89) n-Butylbenzene          | 13.940         | 91   | 11610      | 0.304   | ug/l   | 92       |
| 93) 1,2,4-Trichlorobenzene  | 15.265         | 180  | 9110       | 0.776   | ug/l   | 96       |
| 95) Naphthalene             | 15.504         | 128  | 42043      | 4.716   | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene  | 15.702         | 180  | 10090      | 2.571   | ug/l   | 96       |
| -----                       |                |      |            |         |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN020222\  
Data File : VN070898.D  
Acq On : 2 Feb 2022 17:43  
Operator : JC/MD  
Sample : N1303-03ME  
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
CC-6ME

Quant Time: Feb 02 18:02:01 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N020222W.M  
Quant Title : SW846 8260  
QLast Update : Wed Feb 02 15:30:39 2022  
Response via : Initial Calibration

