

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070900.D
 Acq On : 2 Feb 2022 18:34
 Operator : JC/MD
 Sample : N1303-05ME
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CC-7ME

Quant Time: Feb 02 18:55:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1133032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1795373	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1433059	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	662135	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	705363	45.992	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.980%
35) Dibromofluoromethane	8.019	113	505258	46.978	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.960%
50) Toluene-d8	10.441	98	2105486	47.679	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.360%
62) 4-Bromofluorobenzene	12.731	95	715790	44.985	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.960%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	11.846	91	58762	1.075	ug/l	92
68) m/p-Xylenes	11.950	106	124345	6.138	ug/l	95
69) o-Xylene	12.280	106	54438	2.685	ug/l	92
73) Isopropylbenzene	12.578	105	30201	0.523	ug/l	97
78) n-propylbenzene	12.919	91	122545	1.860	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	349449	7.462	ug/l	95
84) 1,2,4-Trimethylbenzene	13.367	105	1172192	25.596	ug/l	97
86) p-Isopropyltoluene	13.613	119	63455	1.408	ug/l	97
95) Naphthalene	15.504	128	297654	11.509	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
Data File : VN070900.D
Acq On : 2 Feb 2022 18:34
Operator : JC/MD
Sample : N1303-05ME
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CC-7ME

Quant Time: Feb 02 18:55:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

