

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
 Data File : VN070915.D
 Acq On : 3 Feb 2022 1:48
 Operator : JC/MD
 Sample : N1342-11DL 20X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31528DL

Quant Time: Feb 03 06:01:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 02 15:30:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	926282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1461074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1321185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	648507	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	597857	47.683	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.360%
35) Dibromofluoromethane	8.021	113	425439	48.607	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.220%
50) Toluene-d8	10.440	98	1807017	50.282	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.560%
62) 4-Bromofluorobenzene	12.731	95	688564	53.175	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.340%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.827	74	124872	17.779	ug/l	82
11) Tert butyl alcohol	5.380	59	267174	101.409	ug/l	100
16) Acetone	4.288	43	511346	68.205	ug/l	98
18) Methyl Acetate	4.857	43	100558	5.465	ug/l	94
19) Methyl tert-butyl Ether	5.621	73	347357	9.746	ug/l #	88
25) 2-Butanone	7.340	43	321014	29.664	ug/l	93
31) Cyclohexane	8.091	56	50334	2.363	ug/l #	60
39) Methylcyclohexane	9.459	83	26673	1.526	ug/l	93
40) Benzene	8.461	78	5219735	116.844	ug/l	99
52) Toluene	10.505	92	4422708	159.785	ug/l	96
67) Ethyl Benzene	11.843	91	352464	6.992	ug/l	95
68) m/p-Xylenes	11.950	106	728449	39.002	ug/l	91
69) o-Xylene	12.280	106	267200	14.294	ug/l	88
80) 1,3,5-Trimethylbenzene	13.058	105	133888	2.919	ug/l	98
84) 1,2,4-Trimethylbenzene	13.366	105	247873	5.526	ug/l	95
95) Naphthalene	15.504	128	41526	4.679	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN020222\
Data File : VN070915.D
Acq On : 3 Feb 2022 1:48
Operator : JC/MD
Sample : N1342-11DL 20X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31528DL

Quant Time: Feb 03 06:01:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N020222W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 02 15:30:39 2022
Response via : Initial Calibration

